

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
Lab Code: CHEM Case No.: P3671 SAS No.: P3671 SDG NO.: P3671
Lab File ID: VY019211.D Date Analyzed: 08/22/2024
Instrument ID: MSVOA_Y Time Analyzed: 11:22
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	77851	13.353				
UPPER LIMIT	155702	13.853				
LOWER LIMIT	38925.5	12.853				
EPA SAMPLE NO.						
S-904-J-SO-0-0.5-081924	57008	13.35				
S-911-K1-SO-0-0.5-081924RE	38786 *	13.35				
S-911-K1-SO-0.5-1-081924RE	25907 *	13.35				
S-911-K1-SO-0-0.5-081924-FDRE	28511 *	13.35				
S-905-KJ-SO-0-0.5-081924-FD	46673	13.35				
S-928-K1-SO-0-0.5-081924RE	28585 *	13.35				
S-928-K1-SO-0.5-1-081924RE	28694 *	13.35				
S-912-J-SO-0-0.5-081924RE	27279 *	13.35				
S-912-J-SO-0.5-1-081924	42316	13.35				
VY0822SBL01	55761	13.35				
VY0822SBS01	76163	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.