

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: JAC005

Lab Code: CHEM Case No.: P3671 SAS No.: P3671 SDG No.: P3671

Instrument ID: MSVOA_Y Calibration Date/Time: 08/22/2024 11:22

Lab File ID: VY019211.D Init. Calib. Date(s): 08/15/2024 08/15/2024

Heated Purge: (Y/N) Y Init. Calib. Time(s): 17:24 19:50

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Bromomethane	50.000	60.60	-21.192	20
Chloroethane	50.000	51.30	-2.603996	20
Methylene Chloride	50.000	51.60	-3.204002	20
Cyclohexane	50.000	52.65	-5.302002	20