



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY081524\
Data File : VY019070.D
Acq On : 15 Aug 2024 16:51
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y081524S.M
Title : SW846 8260
Last Update : Fri Aug 16 05:10:30 2024



AutoFind: Scans 1751, 1752, 1753; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	8824	PASS
75	95	30	60	55.6	23968	PASS
95	95	100	100	100.0	43144	PASS
96	95	5	9	7.0	3035	PASS
173	174	0.00	2	1.4	477	PASS
174	95	50	100	81.5	35173	PASS
175	174	5	9	8.2	2881	PASS
176	174	95	101	99.2	34888	PASS
177	176	5	9	7.3	2561	PASS