

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083142.D
 Acq On : 07 Aug 2024 14:49
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN080724

Quant Time: Aug 08 06:35:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	109	0.00
2 T	Dichlorodifluoromethane	0.567	0.534	5.8	102	0.00
3 P	Chloromethane	0.581	0.537	7.6	99	0.00
4 C	Vinyl Chloride	0.592	0.544	8.1#	99	0.00
5 T	Bromomethane	0.368	0.316	14.1	97	0.00
6 T	Chloroethane	0.371	0.324	12.7	99	0.00
7 T	Trichlorofluoromethane	0.979	0.922	5.8	101	0.00
8 T	Diethyl Ether	0.364	0.337	7.4	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.539	0.509	5.6	101	0.00
10 T	Methyl Iodide	0.710	0.672	5.4	99	0.00
11 T	Tert butyl alcohol	0.148	0.123	16.9	92	0.00
12 CM	1,1-Dichloroethene	0.554	0.505	8.8#	101	0.00
13 T	Acrolein	0.096	0.084	12.5	90	0.00
14 T	Allyl chloride	1.047	0.862	17.7	101	0.00
15 T	Acrylonitrile	0.304	0.265	12.8	95	0.00
16 T	Acetone	0.278	0.235	15.5	90	0.00
17 T	Carbon Disulfide	1.622	1.417	12.6	99	0.00
18 T	Methyl Acetate	0.830	0.704	15.2	99	0.00
19 T	Methyl tert-butyl Ether	2.001	1.872	6.4	101	0.00
20 T	Methylene Chloride	0.641	0.559	12.8	102	0.00
21 T	trans-1,2-Dichloroethene	0.573	0.523	8.7	99	-0.01
22 T	Diisopropyl ether	1.969	1.846	6.2	100	0.00
23 T	Vinyl Acetate	2.018	1.900	5.8	99	0.00
24 P	1,1-Dichloroethane	1.073	1.002	6.6	101	0.00
25 T	2-Butanone	0.428	0.362	15.4	93	0.00
26 T	2,2-Dichloropropane	0.997	0.954	4.3	102	0.00
27 T	cis-1,2-Dichloroethene	0.691	0.638	7.7	101	0.00
28 T	Bromochloromethane	0.439	0.403	8.2	100	0.00
29 T	Tetrahydrofuran	0.276	0.246	10.9	96	0.00
30 C	Chloroform	1.115	1.057	5.2#	101	0.00
31 T	Cyclohexane	1.055	0.917	13.1	102	0.00
32 T	1,1,1-Trichloroethane	1.055	0.995	5.7	101	0.00
33 S	1,2-Dichloroethane-d4	0.712	0.720	-1.1	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	109	0.00
35 S	Dibromofluoromethane	0.312	0.322	-3.2	106	0.00
36 T	1,1-Dichloropropene	0.472	0.451	4.4	103	0.00
37 T	Ethyl Acetate	0.529	0.456	13.8	100	0.00
38 T	Carbon Tetrachloride	0.532	0.509	4.3	102	0.00
39 T	Methylcyclohexane	0.580	0.555	4.3	101	0.00
40 TM	Benzene	1.406	1.340	4.7	101	0.00
41 T	Methacrylonitrile	0.301	0.259	14.0	99	0.00
42 TM	1,2-Dichloroethane	0.512	0.485	5.3	102	0.00
43 T	Isopropyl Acetate	1.211	0.841	30.6#	100	0.00
44 TM	Trichloroethene	0.335	0.321	4.2	102	0.00
45 C	1,2-Dichloropropane	0.334	0.322	3.6#	101	0.00
46 T	Dibromomethane	0.239	0.225	5.9	98	0.00
47 T	Bromodichloromethane	0.537	0.505	6.0	102	0.00
48 T	Methyl methacrylate	0.437	0.390	10.8	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
 Data File : VN083142.D
 Acq On : 07 Aug 2024 14:49
 Operator : JC\MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICSVN080724

Quant Time: Aug 08 06:35:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 08 06:30:41 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	93	0.00
50 S	Toluene-d8	1.164	1.215	-4.4	105	0.00
51 T	4-Methyl-2-Pentanone	0.500	0.462	7.6	97	0.00
52 CM	Toluene	0.889	0.862	3.0#	101	0.00
53 T	t-1,3-Dichloropropene	0.551	0.549	0.4	100	0.00
54 T	cis-1,3-Dichloropropene	0.586	0.568	3.1	102	0.00
55 T	1,1,2-Trichloroethane	0.318	0.308	3.1	101	0.00
56 T	Ethyl methacrylate	0.600	0.569	5.2	99	0.00
57 T	1,3-Dichloropropane	0.568	0.538	5.3	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.254	0.249	2.0	104	0.00
59 T	2-Hexanone	0.387	0.348	10.1	96	0.00
60 T	Dibromochloromethane	0.385	0.381	1.0	100	0.00
61 T	1,2-Dibromoethane	0.335	0.315	6.0	99	0.00
62 S	4-Bromofluorobenzene	0.454	0.477	-5.1	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
64 T	Tetrachloroethene	0.331	0.319	3.6	102	0.00
65 PM	Chlorobenzene	1.105	1.050	5.0	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.390	0.365	6.4	100	0.00
67 C	Ethyl Benzene	2.027	1.936	4.5#	102	0.00
68 T	m/p-Xylenes	0.759	0.732	3.6	102	0.00
69 T	o-Xylene	0.749	0.699	6.7	99	0.00
70 T	Styrene	1.258	1.220	3.0	100	0.00
71 P	Bromoform	0.295	0.282	4.4	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
73 T	Isopropylbenzene	4.182	3.922	6.2	101	0.00
74 T	N-amyl acetate	2.046	1.790	12.5	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.183	1.018	13.9	96	0.00
76 T	1,2,3-Trichloropropane	1.097	0.924	15.8	95	0.00
77 T	Bromobenzene	0.929	0.867	6.7	99	0.00
78 T	n-propylbenzene	4.816	4.592	4.7	101	0.00
79 T	2-Chlorotoluene	3.053	2.833	7.2	101	0.00
80 T	1,3,5-Trimethylbenzene	3.501	3.338	4.7	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.504	0.475	5.8	105	0.00
82 T	4-Chlorotoluene	3.062	2.849	7.0	101	0.00
83 T	tert-Butylbenzene	3.101	2.863	7.7	99	0.00
84 T	1,2,4-Trimethylbenzene	3.529	3.281	7.0	98	0.00
85 T	sec-Butylbenzene	4.231	3.979	6.0	99	0.00
86 T	p-Isopropyltoluene	3.493	3.323	4.9	99	0.00
87 T	1,3-Dichlorobenzene	1.748	1.607	8.1	99	0.00
88 T	1,4-Dichlorobenzene	1.762	1.604	9.0	99	0.00
89 T	n-Butylbenzene	3.027	2.935	3.0	100	0.00
90 T	Hexachloroethane	0.675	0.631	6.5	100	0.00
91 T	1,2-Dichlorobenzene	1.692	1.531	9.5	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.287	0.240	16.4	93	0.00
93 T	1,2,4-Trichlorobenzene	0.948	0.873	7.9	97	0.00
94 T	Hexachlorobutadiene	0.421	0.365	13.3	98	0.00
95 T	Naphthalene	3.357	3.001	10.6	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN080724\
Data File : VN083142.D
Acq On : 07 Aug 2024 14:49
Operator : JC\MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN080724

Quant Time: Aug 08 06:35:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N080724W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 08 06:30:41 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.938	0.848	9.6	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6