

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3671 SAS No.: P3671 SDG NO.: P3671
 Lab File ID: VN083424.D Date Analyzed: 08/22/2024
 Instrument ID: MSVOA_N Time Analyzed: 08:46
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	120986	13.794				
UPPER LIMIT	241972	14.294				
LOWER LIMIT	60493	13.294				
EPA SAMPLE NO.						
TB-01-081924	136526	13.79				
VN0822WBL01	145694	13.79				
VN0822WBS01	121885	13.79				
VN0822WBSD01	107486	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.