

Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 82N080724W.M
 Title : SW846 8260
 Last Update : Thu Aug 08 06:30:41 2024
 Response Via : Initial Calibration

Calibration Files

1 =VN083140.D 5 =VN083139.D 10 =VN083138.D 20 =VN083137.D 50 =VN083136.D 100 =VN083135.D

	Compound	1	5	10	20	50	100	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.542	0.597	0.579	0.566	0.571	0.548	0.567	3.55
3) P	Chloromethane	0.576	0.595	0.589	0.584	0.591	0.547	0.581	3.06
4) C	Vinyl Chloride	0.607	0.586	0.612	0.588	0.598	0.562	0.592	3.02#
5) T	Bromomethane		0.413	0.385	0.370	0.357	0.313	0.368	10.07
6) T	Chloroethane	0.436	0.381	0.355	0.356	0.356	0.340	0.371	9.33
7) T	Trichlorofluor...	0.901	1.003	1.013	1.005	0.995	0.955	0.979	4.42
8) T	Diethyl Ether	0.338	0.380	0.385	0.367	0.366	0.348	0.364	4.97
9) T	1,1,2-Trichlor...	0.510	0.558	0.550	0.548	0.548	0.524	0.539	3.41
10) T	Methyl Iodide		0.688	0.710	0.703	0.740	0.707	0.710	2.72
11) T	Tert butyl alc...		0.155	0.155	0.146	0.147	0.136	0.148	5.22
12) CM	1,1-Dichloroet...	0.575	0.581	0.564	0.538	0.544	0.524	0.554	4.08#
13) T	Acrolein		0.095	0.101	0.083	0.102	0.100	0.096	7.99
14) T	Allyl chloride	1.210	1.099	1.083	1.063	0.933	0.896	1.047	11.00
15) T	Acrylonitrile	0.306	0.311	0.306	0.304	0.306	0.293	0.304	1.95
16) T	Acetone	0.265	0.274	0.284	0.260	0.286	0.302	0.278	5.51
17) T	Carbon Disulfide	1.879	1.630	1.599	1.596	1.560	1.466	1.622	8.52
18) T	Methyl Acetate	0.985	0.844	0.842	0.773	0.776	0.758	0.830	10.20
19) T	Methyl tert-bu...	1.953	2.053	2.015	2.028	2.028	1.927	2.001	2.46
20) T	Methylene Chlo...	0.805	0.640	0.614	0.619	0.599	0.569	0.641	13.03
21) T	trans-1,2-Dich...	0.599	0.568	0.586	0.577	0.574	0.533	0.573	3.93
22) T	Diisopropyl ether	1.945	1.966	2.009	1.976	2.008	1.909	1.969	1.94
23) T	Vinyl Acetate	1.930	2.006	2.040	2.046	2.096	1.990	2.018	2.80
24) P	1,1-Dichloroet...	1.035	1.093	1.096	1.093	1.084	1.038	1.073	2.65
25) T	2-Butanone	0.453	0.414	0.430	0.424	0.427	0.418	0.428	3.22
26) T	2,2-Dichloropr...	0.925	0.984	1.032	1.019	1.023	0.997	0.997	3.95
27) T	cis-1,2-Dichlo...	0.721	0.687	0.701	0.691	0.689	0.658	0.691	2.99
28) T	Bromochloromet...	0.465	0.489	0.391	0.420	0.439	0.427	0.439	7.91
29) T	Tetrahydrofuran	0.271	0.280	0.286	0.272	0.279	0.270	0.276	2.24
30) C	Chloroform	1.102	1.122	1.132	1.110	1.137	1.087	1.115	1.71#
31) T	Cyclohexane		1.238	1.110	1.010	0.982	0.933	1.055	11.50
32) T	1,1,1-Trichlor...	1.007	1.081	1.076	1.066	1.075	1.027	1.055	2.89
33) S	1,2-Dichloroet...		0.786	0.710	0.601	0.745	0.717	0.712	9.68
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...		0.346	0.307	0.264	0.332	0.311	0.312	9.99
36) T	1,1-Dichloropr...	0.488	0.463	0.477	0.472	0.475	0.457	0.472	2.33
37) T	Ethyl Acetate	0.665	0.515	0.512	0.502	0.499	0.479	0.529	12.83
38) T	Carbon Tetrach...	0.493	0.539	0.541	0.548	0.546	0.524	0.532	3.93
39) T	Methylcyclohexane	0.572	0.575	0.583	0.587	0.597	0.565	0.580	1.96
40) TM	Benzene	1.343	1.403	1.429	1.431	1.447	1.386	1.406	2.68
41) T	Methacrylonitrile	0.377	0.290	0.277	0.300	0.283	0.276	0.301	12.78
42) TM	1,2-Dichloroet...	0.498	0.524	0.524	0.514	0.520	0.494	0.512	2.63
43) T	Isopropyl Acetate	2.156	1.289	1.075	0.956	0.921	0.869	1.211	40.18
44) TM	Trichloroethene	0.310	0.340	0.341	0.343	0.344	0.331	0.335	3.94
45) C	1,2-Dichloropr...	0.316	0.326	0.341	0.344	0.346	0.331	0.334	3.50#
46) T	Dibromomethane	0.218	0.243	0.240	0.245	0.251	0.238	0.239	4.73
47) T	Bromodichlorom...	0.544	0.541	0.533	0.540	0.539	0.521	0.537	1.55
48) T	Methyl methacr...	0.506	0.423	0.432	0.421	0.425	0.415	0.437	7.88
49) T	1,4-Dioxane	0.008	0.008	0.008	0.008	0.008	0.008	0.008	2.64
50) S	Toluene-d8		1.240	1.133	0.985	1.261	1.201	1.164	9.56
51) T	4-Methyl-2-Pen...	0.457	0.499	0.516	0.516	0.517	0.495	0.500	4.62
52) CM	Toluene	0.824	0.885	0.900	0.905	0.926	0.892	0.889	3.91#
53) T	t-1,3-Dichloro...	0.469	0.546	0.552	0.573	0.597	0.569	0.551	7.99
54) T	cis-1,3-Dichlo...	0.546	0.588	0.595	0.597	0.606	0.586	0.586	3.54
55) T	1,1,2-Trichlor...	0.283	0.314	0.327	0.331	0.334	0.322	0.318	5.95
56) T	Ethyl methacry...	0.553	0.594	0.612	0.617	0.624	0.601	0.600	4.24


Method Path : Z:\voasrv\HPCHEM1\MSVOA_N\methods\
 Method File : 82N080724W.M

57)	T	1,3-Dichloropr...	0.552	0.545	0.575	0.581	0.592	0.561	0.568	3.17
58)	T	2-Chloroethyl ...	0.243	0.250	0.262	0.242	0.261	0.264	0.254	3.91
59)	T	2-Hexanone	0.367	0.385	0.389	0.395	0.397	0.388	0.387	2.70
60)	T	Dibromochlorom...	0.329	0.383	0.388	0.397	0.413	0.400	0.385	7.68
61)	T	1,2-Dibromoethane	0.317	0.332	0.342	0.336	0.347	0.334	0.335	3.13
62)	S	4-Bromofluorob...	0.483	0.432	0.390	0.492	0.473	0.454		9.36
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.320	0.324	0.343	0.338	0.343	0.319	0.331	3.46
65)	PM	Chlorobenzene	1.069	1.094	1.153	1.101	1.142	1.070	1.105	3.23
66)	T	1,1,1,2-Tetrac...	0.370	0.401	0.403	0.385	0.399	0.380	0.390	3.37
67)	C	Ethyl Benzene	1.957	2.032	2.070	2.061	2.075	1.967	2.027	2.59#
68)	T	m/p-Xylenes	0.706	0.759	0.785	0.771	0.787	0.747	0.759	3.99
69)	T	o-Xylene	0.703	0.750	0.773	0.759	0.773	0.735	0.749	3.60
70)	T	Styrene	1.100	1.261	1.300	1.288	1.336	1.261	1.258	6.51
71)	P	Bromoform	0.267	0.275	0.292	0.309	0.318	0.310	0.295	7.01
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	4.254	4.249	4.386	4.133	4.249	3.822	4.182	4.63
74)	T	N-amyl acetate	2.341	2.051	2.108	1.945	2.003	1.826	2.046	8.50
75)	P	1,1,2,2-Tetrac...	1.234	1.237	1.210	1.163	1.167	1.085	1.183	4.85
76)	T	1,2,3-Trichlor...	1.233	1.083	1.128	1.094	1.063	0.984	1.097	7.47
77)	T	Bromobenzene	0.900	0.975	0.946	0.927	0.959	0.868	0.929	4.27
78)	T	n-propylbenzene	4.800	4.849	4.998	4.753	4.968	4.526	4.816	3.55
79)	T	2-Chlorotoluene	3.146	3.083	3.184	3.027	3.076	2.802	3.053	4.42
80)	T	1,3,5-Trimethy...	3.479	3.563	3.641	3.490	3.589	3.245	3.501	3.98
81)	T	trans-1,4-Dich...	0.530	0.468	0.524	0.493	0.508	0.504		4.97
82)	T	4-Chlorotoluene	3.261	3.081	3.116	3.014	3.104	2.799	3.062	4.99
83)	T	tert-Butylbenzene	3.103	3.199	3.192	3.098	3.162	2.850	3.101	4.19
84)	T	1,2,4-Trimethy...	3.416	3.592	3.679	3.526	3.654	3.305	3.529	4.10
85)	T	sec-Butylbenzene	4.235	4.265	4.325	4.202	4.384	3.976	4.231	3.33
86)	T	p-Isopropyltol...	3.355	3.569	3.510	3.491	3.682	3.354	3.493	3.62
87)	T	1,3-Dichlorobe...	1.870	1.708	1.757	1.739	1.771	1.643	1.748	4.29
88)	T	1,4-Dichlorobe...	1.862	1.795	1.751	1.728	1.783	1.655	1.762	3.96
89)	T	n-Butylbenzene	2.851	3.017	3.045	3.023	3.216	3.011	3.027	3.84
90)	T	Hexachloroethane	0.620	0.696	0.708	0.686	0.689	0.650	0.675	4.92
91)	T	1,2-Dichlorobe...	1.741	1.731	1.739	1.664	1.704	1.571	1.692	3.90
92)	T	1,2-Dibromo-3-...	0.310	0.307	0.289	0.277	0.281	0.258	0.287	6.84
93)	T	1,2,4-Trichlor...	1.014	0.893	0.938	0.943	0.984	0.914	0.948	4.72
94)	T	Hexachlorobuta...	0.480	0.445	0.429	0.397	0.409	0.368	0.421	9.32
95)	T	Naphthalene	3.429	3.230	3.435	3.365	3.481	3.202	3.357	3.44
96)	T	1,2,3-Trichlor...	0.972	0.920	0.974	0.920	0.963	0.877	0.938	4.09

(#) = Out of Range

Isopropyl Acetate

Response Ratio



$R = -3.250e-002 A^2 + 9.200e-001 A + 2.928e-002$
 Coef of Det (r^2) = 0.999905 Curve Fit: Quadratic
 Method Name: Z:\voasrv\HPCHEM1\MSVOA N\methods\82N080724W.M
 Calibration Table Last Updated: Thu Aug 08 06:30:41 2024