



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: P3671
Client: JACOBS Engineering Group, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID: 0

Total Voc :
Total Concentration:

A
B
C
D



SAMPLE DATA

A

B

C

D

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/19/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	08/19/24
Client Sample ID:	TB-01-081924	SDG No.:	P3671
Lab Sample ID:	P3671-15	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group6
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN083436.D	1		08/22/24 15:20	VN082224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	1.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	1.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	1.00	ug/L
67-64-1	Acetone	1.40	U	1.40	5.00	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	1.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	1.00	ug/L
71-43-2	Benzene	0.16	U	0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	1.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	1.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	1.00	ug/L
108-88-3	Toluene	0.18	U	0.18	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	1.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	1.00	ug/L
100-41-4	Ethyl Benzene	0.16	U	0.16	1.00	ug/L

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	08/19/24
Project:	Former Schlumberger Site Princeton NJ	Date Received:	08/19/24
Client Sample ID:	TB-01-081924	SDG No.:	P3671
Lab Sample ID:	P3671-15	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group6
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN083436.D	1		08/22/24 15:20	VN082224

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
179601-23-1	m/p-Xylenes	0.31	U	0.31	2.00	ug/L
1330-20-7	Total Xylenes	0.45	U	0.45	3.00	ug/L
95-47-6	o-Xylene	0.14	U	0.14	1.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	59.1		70 (74) - 130 (125)	118%	SPK: 50
1868-53-7	Dibromofluoromethane	52.0		70 (75) - 130 (124)	104%	SPK: 50
2037-26-5	Toluene-d8	50.0		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	57.6		70 (77) - 130 (121)	115%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	128000	8.224			
540-36-3	1,4-Difluorobenzene	263000	9.1			
3114-55-4	Chlorobenzene-d5	286000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	137000	13.794			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	P3671	OrderDate:	8/20/2024 10:36:00 AM
Client:	JACOBS Engineering Group, Inc.	Project:	Former Schlumberger Site Princeton NJ
Contact:	Mary I. Murphy	Location:	G11,VOA Ref. #2 Soil,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
P3671-01	S-904-J-SO-0-0.5-081 924	SOIL	VOCMS Group5	8260D	08/19/24		08/22/24	08/19/24
P3671-02	S-904-J-SO-0.5-1.0-0 81924	SOIL	VOCMS Group5	8260D	08/19/24		08/22/24	08/19/24
P3671-03	S-911-K1-SO-0-0.5-08 1924	SOIL	VOCMS Group5	8260D	08/19/24		08/22/24	08/19/24
P3671-03RE	S-911-K1-SO-0-0.5-08 1924RE	SOIL	VOCMS Group5	8260D	08/19/24		08/22/24	08/19/24
P3671-04	S-911-K1-SO-0.5-1-08 1924	SOIL	VOCMS Group5	8260D	08/19/24		08/22/24	08/19/24
P3671-04RE	S-911-K1-SO-0.5-1-08 1924RE	SOIL	VOCMS Group5	8260D	08/19/24		08/22/24	08/19/24
P3671-05	S-911-K1-SO-0-0.5-08 1924-FD	SOIL	VOCMS Group5	8260D	08/19/24		08/22/24	08/19/24
P3671-05RE	S-911-K1-SO-0-0.5-08 1924-FDRE	SOIL	VOCMS Group5	8260D	08/19/24		08/22/24	08/19/24
P3671-06	S-905-KJ-SO-0-0.5-08 1924	SOIL	VOCMS Group5	8260D	08/19/24		08/22/24	08/19/24

LAB CHRONICLE

P3671-07	S-905-KJ-SO-0-0.5-08 1924-FD	SOIL		08/19/24	08/19/24
			VOCMS Group5	8260D	08/22/24
P3671-08	S-905-J-SO-0.5-1.0-0 81924	SOIL		08/19/24	08/19/24
			VOCMS Group5	8260D	08/22/24
P3671-09	S-928-K1-SO-0-0.5-08 1924	SOIL		08/19/24	08/19/24
			VOCMS Group5	8260D	08/22/24
P3671-09RE	S-928-K1-SO-0-0.5-08 1924RE	SOIL		08/19/24	08/19/24
			VOCMS Group5	8260D	08/22/24
P3671-10	S-928-K1-SO-0.5-1-08 1924	SOIL		08/19/24	08/19/24
			VOCMS Group5	8260D	08/22/24
P3671-10RE	S-928-K1-SO-0.5-1-08 1924RE	SOIL		08/19/24	08/19/24
			VOCMS Group5	8260D	08/22/24
P3671-11	S-903-J-SO-0-0.5-081 924	SOIL		08/19/24	08/19/24
			VOCMS Group5	8260D	08/22/24
P3671-12	S-903-J-SO-0.5-1.0-0 81924	SOIL		08/19/24	08/19/24
			VOCMS Group5	8260D	08/22/24
P3671-13	S-912-J-SO-0-0.5-081 924	SOIL		08/19/24	08/19/24
			VOCMS Group5	8260D	08/22/24
P3671-13RE	S-912-J-SO-0-0.5-081 924RE	SOIL		08/19/24	08/19/24
			VOCMS Group5	8260D	08/22/24
P3671-14	S-912-J-SO-0.5-1-081 924	SOIL		08/19/24	08/19/24
			VOCMS Group5	8260D	08/22/24

LAB CHRONICLE

P3671-15	TB-01-081924	Water	VOCMS Group6	8260-Low	08/19/24	08/22/24	08/19/24
----------	--------------	-------	--------------	----------	----------	----------	----------