



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

**GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: P3845

MATRIX: Water

METHOD: 8151A/3510

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified.			✓
2. Standard Summary Submitted.			✓
3. Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis, 12 HOURS IF 8000 SERIES METHOD.  The Initial Calibration met the requirements .  The Continuous Calibration met the requirements			✓
4. Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
5. Surrogate Recoveries Meet Criteria  If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			✓
6. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria  If not met, list those compounds and their recoveries which fall outside the acceptable range.  The Blank Spike met requirements for all samples .  The Blank Spike Duplicate met requirements for all samples .  The RPD met criteria .			✓
7. Retention Time Shift Meet Criteria (if applicable)  Comments:			✓
8. Extraction Holding Time Met  If not met, list number of days exceeded for each sample:			✓
9. Analysis Holding Time Met  If not met, list those compounds and their recoveries which fall outside the acceptable range.			✓



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

**GC ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)**

NA NO YES

ADDITIONAL COMMENTS:

\_\_\_\_\_  
QA REVIEW

\_\_\_\_\_  
Date