

WATER DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE

Lab Name: Chemtech **Client:** Chemtech Consulting Group
Lab Code: CHEM **Cas No:** P3845 **SAS No :** P3845 **SDG No:** P3845
Matrix Spike - EPA Sample No : PB163245BS **Datafile:** FF014474.D

COMPOUND	SPIKE ADDED ug/L	CONCENTRATION ug/L	LCS/LCSD CONCENTRATION ug/L	% REC	QC LIMITS
DRO	200	0	194	97	78-117