

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

Continuing Calib Date: 09/10/2024 Initial Calibration Date(s): 09/03/2024 09/03/2024

Continuing Calib Time: 14:18 Initial Calibration Time(s): 13:26 15:03

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.33	7.33	7.23	7.43	0.00
MCP	7.51	7.51	7.41	7.61	0.00
2,4-DCAA	7.15	7.15	7.05	7.25	0.00
Dalapon	2.59	2.59	2.49	2.69	0.00
MCPA	7.65	7.65	7.55	7.75	0.00
DICHLORPROP	8.02	8.02	7.92	8.12	0.00
2,4-D	8.24	8.24	8.14	8.34	0.00
2,4,5-TP(Silvex)	9.10	9.10	9.00	9.20	0.00
2,4,5-T	9.39	9.39	9.29	9.49	0.00
2,4-DB	9.95	9.96	9.86	10.06	0.01
Dinoseb	11.13	11.14	11.04	11.24	0.01
Pentachlorophenol	8.53	8.53	8.43	8.63	0.00
4-Nitrophenol	6.94	6.95	6.85	7.05	0.01
PICLORAM	10.95	10.96	10.86	11.06	0.01
DCPA	11.43	11.44	11.34	11.54	0.01
3,5-DICHLOROBENZ	6.33	6.33	6.23	6.43	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

Continuing Calib Date: 09/10/2024 Initial Calibration Date(s): 09/03/2024 09/03/2024

Continuing Calib Time: 14:18 Initial Calibration Time(s): 13:26 15:03

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.93	7.93	7.83	8.03	0.00
MCPD	8.03	8.03	7.93	8.13	0.00
2,4-DCAA	7.73	7.73	7.63	7.83	0.00
Dalapon	2.69	2.69	2.59	2.79	0.00
MCPA	8.27	8.27	8.17	8.37	0.00
DICHLORPROP	8.64	8.64	8.54	8.74	0.00
2,4-D	8.97	8.97	8.87	9.07	0.00
2,4,5-TP(Silvex)	9.87	9.87	9.77	9.97	0.00
2,4,5-T	10.29	10.28	10.18	10.38	0.00
2,4-DB	10.85	10.85	10.75	10.95	0.00
Dinoseb	11.23	11.23	11.13	11.33	0.00
Pentachlorophenol	9.49	9.49	9.39	9.59	0.00
4-Nitrophenol	7.26	7.26	7.16	7.36	0.00
PICLORAM	12.32	12.32	12.22	12.42	0.00
DCPA	12.27	12.27	12.17	12.37	0.00
3,5-DICHLOROBENZ	6.69	6.69	6.59	6.79	0.00