

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

Continuing Calib Date: 09/13/2024 Initial Calibration Date(s): 09/12/2024 09/12/2024

Continuing Calib Time: 01:22 Initial Calibration Time(s): 21:24 22:59

GC Column: RTX-CLP ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.29	7.29	7.19	7.39	0.00
MCP	7.47	7.47	7.37	7.57	0.00
2,4-DCAA	7.11	7.11	7.01	7.21	0.00
Dalapon	2.56	2.55	2.45	2.65	-0.01
MCPA	7.62	7.62	7.52	7.72	0.01
DICHLORPROP	7.98	7.98	7.88	8.08	0.00
2,4-D	8.20	8.20	8.10	8.30	0.00
2,4,5-TP(Silvex)	9.06	9.06	8.96	9.16	0.00
2,4,5-T	9.34	9.34	9.24	9.44	0.00
2,4-DB	9.91	9.91	9.81	10.01	0.01
Dinoseb	11.08	11.08	10.98	11.18	0.00
Pentachlorophenol	8.49	8.49	8.39	8.59	0.00
4-Nitrophenol	6.91	6.91	6.81	7.01	0.00
PICLORAM	10.90	10.90	10.80	11.00	0.00
DCPA	11.38	11.38	11.28	11.48	0.00
3,5-DICHLOROBENZ	6.30	6.30	6.20	6.40	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

Continuing Calib Date: 09/13/2024 Initial Calibration Date(s): 09/12/2024 09/12/2024

Continuing Calib Time: 01:22 Initial Calibration Time(s): 21:24 22:59

GC Column: RTX-CLP2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DICAMBA	7.83	7.83	7.73	7.93	0.00
MCP	7.93	7.93	7.83	8.03	0.00
2,4-DCAA	7.63	7.63	7.53	7.73	0.00
Dalapon	2.62	2.62	2.52	2.72	0.00
MCPA	8.17	8.17	8.07	8.27	0.00
DICHLORPROP	8.54	8.54	8.44	8.64	0.01
2,4-D	8.86	8.86	8.76	8.96	0.00
2,4,5-TP(Silvex)	9.76	9.76	9.66	9.86	0.00
2,4,5-T	10.17	10.17	10.07	10.27	0.00
2,4-DB	10.73	10.73	10.63	10.83	0.00
Dinoseb	11.11	11.11	11.01	11.21	0.00
Pentachlorophenol	9.38	9.38	9.28	9.48	0.00
4-Nitrophenol	7.17	7.17	7.07	7.27	0.00
PICLORAM	12.19	12.19	12.09	12.29	0.00
DCPA	12.15	12.15	12.05	12.25	0.00
3,5-DICHLOROBENZ	6.60	6.60	6.50	6.70	0.00