

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 09/03/2024 09/03/2024

Client Sample No.: CCAL01 Date Analyzed: 09/10/2024

Lab Sample No.: HSTDCCC750 Data File : PS027618.D Time Analyzed: 14:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.386	9.289	9.489	720.670	712.500	1.1
2,4,5-TP(Silvex)	9.098	9.002	9.202	715.380	712.500	0.4
2,4-D	8.240	8.143	8.343	695.580	705.000	-1.3
2,4-DB	9.951	9.855	10.055	693.470	712.500	-2.7
2,4-DCAA	7.146	7.049	7.249	740.100	750.000	-1.3
3,5-DICHLOROBENZOIC ACID	6.331	6.233	6.433	697.050	697.500	-0.1
4-Nitrophenol	6.943	6.846	7.046	698.710	682.500	2.4
Dalapon	2.586	2.488	2.688	699.320	682.500	2.5
DCPA	11.433	11.337	11.537	731.730	720.000	1.6
DICAMBA	7.327	7.229	7.429	708.340	705.000	0.5
DICHLORPROP	8.016	7.919	8.119	688.250	705.000	-2.4
Dinoseb	11.134	11.037	11.237	780.880	705.000	10.8
MCPA	7.650	7.552	7.752	68.030	69.750	-2.5
MCPP	7.505	7.407	7.607	71.590	70.500	1.5
Pentachlorophenol	8.529	8.432	8.632	736.060	712.500	3.3
PICLORAM	10.952	10.855	11.055	679.900	712.500	-4.6

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/03/2024 09/03/2024

Client Sample No.: CCAL01 Date Analyzed: 09/10/2024

Lab Sample No.: HSTDCCC750 Data File : PS027618.D Time Analyzed: 14:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.285	10.183	10.383	715.170	712.500	0.4
2,4,5-TP(Silvex)	9.867	9.766	9.966	720.490	712.500	1.1
2,4-D	8.967	8.866	9.066	706.800	705.000	0.3
2,4-DB	10.849	10.748	10.948	706.720	712.500	-0.8
2,4-DCAA	7.730	7.630	7.830	747.480	750.000	-0.3
3,5-DICHLOROBENZOIC ACID	6.687	6.587	6.787	700.850	697.500	0.5
4-Nitrophenol	7.260	7.159	7.359	709.770	682.500	4.0
Dalapon	2.689	2.591	2.791	685.490	682.500	0.4
DCPA	12.267	12.166	12.366	732.200	720.000	1.7
DICAMBA	7.928	7.827	8.027	715.030	705.000	1.4
DICHLORPROP	8.639	8.538	8.738	707.750	705.000	0.4
Dinoseb	11.227	11.126	11.326	797.750	705.000	13.2
MCPA	8.272	8.172	8.372	66.830	69.750	-4.2
MCPP	8.030	7.929	8.129	70.210	70.500	-0.4
Pentachlorophenol	9.493	9.392	9.592	728.720	712.500	2.3
PICLORAM	12.317	12.216	12.416	686.060	712.500	-3.7