

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 09/03/2024 09/03/2024

Client Sample No.: CCAL02 Date Analyzed: 09/10/2024

Lab Sample No.: HSTDCCC750 Data File : PS027630.D Time Analyzed: 22:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.383	9.289	9.489	700.120	712.500	-1.7
2,4,5-TP(Silvex)	9.097	9.002	9.202	719.350	712.500	1.0
2,4-D	8.239	8.143	8.343	704.070	705.000	-0.1
2,4-DB	9.950	9.855	10.055	715.240	712.500	0.4
2,4-DCAA	7.145	7.049	7.249	687.350	750.000	-8.4
3,5-DICHLOROBENZOIC ACID	6.331	6.233	6.433	692.970	697.500	-0.6
4-Nitrophenol	6.942	6.846	7.046	696.710	682.500	2.1
Dalapon	2.584	2.488	2.688	689.510	682.500	1.0
DCPA	11.431	11.337	11.537	728.090	720.000	1.1
DICAMBA	7.326	7.229	7.429	710.470	705.000	0.8
DICHLORPROP	8.014	7.919	8.119	678.790	705.000	-3.7
Dinoseb	11.132	11.037	11.237	732.780	705.000	3.9
MCPA	7.649	7.552	7.752	68.650	69.750	-1.6
MCPP	7.504	7.407	7.607	71.770	70.500	1.8
Pentachlorophenol	8.528	8.432	8.632	734.270	712.500	3.1
PICLORAM	10.950	10.855	11.055	728.980	712.500	2.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/03/2024 09/03/2024

Client Sample No.: CCAL02 Date Analyzed: 09/10/2024

Lab Sample No.: HSTDCCC750 Data File : PS027630.D Time Analyzed: 22:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.285	10.183	10.383	722.940	712.500	1.5
2,4,5-TP(Silvex)	9.867	9.766	9.966	720.290	712.500	1.1
2,4-D	8.967	8.866	9.066	708.710	705.000	0.5
2,4-DB	10.849	10.748	10.948	714.910	712.500	0.3
2,4-DCAA	7.730	7.630	7.830	750.060	750.000	0.0
3,5-DICHLOROBENZOIC ACID	6.688	6.587	6.787	698.570	697.500	0.2
4-Nitrophenol	7.260	7.159	7.359	703.630	682.500	3.1
Dalapon	2.690	2.591	2.791	679.040	682.500	-0.5
DCPA	12.267	12.166	12.366	727.660	720.000	1.1
DICAMBA	7.928	7.827	8.027	711.150	705.000	0.9
DICHLORPROP	8.639	8.538	8.738	702.120	705.000	-0.4
Dinoseb	11.227	11.126	11.326	785.670	705.000	11.4
MCPA	8.273	8.172	8.372	68.520	69.750	-1.8
MCPP	8.030	7.929	8.129	70.880	70.500	0.5
Pentachlorophenol	9.493	9.392	9.592	727.040	712.500	2.0
PICLORAM	12.317	12.216	12.416	732.020	712.500	2.7