

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 09/12/2024 09/12/2024

Client Sample No.: CCAL03 Date Analyzed: 09/13/2024

Lab Sample No.: HSTDCCC750 Data File : PS027660.D Time Analyzed: 00:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.340	9.242	9.442	692.280	712.500	-2.8
2,4,5-TP(Silvex)	9.055	8.956	9.156	694.280	712.500	-2.6
2,4-D	8.201	8.101	8.301	675.970	705.000	-4.1
2,4-DB	9.904	9.805	10.005	692.180	712.500	-2.9
2,4-DCAA	7.111	7.012	7.212	727.710	750.000	-3.0
3,5-DICHLOROBENZOIC ACID	6.301	6.201	6.401	677.750	697.500	-2.8
4-Nitrophenol	6.906	6.807	7.007	648.620	682.500	-5.0
Dalapon	2.554	2.454	2.654	654.130	682.500	-4.2
DCPA	11.383	11.283	11.483	700.770	720.000	-2.7
DICAMBA	7.290	7.191	7.391	687.830	705.000	-2.4
DICHLORPROP	7.977	7.878	8.078	679.880	705.000	-3.6
Dinoseb	11.084	10.984	11.184	669.160	705.000	-5.1
MCPA	7.615	7.516	7.716	68.650	69.750	-1.6
MCPP	7.470	7.371	7.571	70.800	70.500	0.4
Pentachlorophenol	8.487	8.388	8.588	696.560	712.500	-2.2
PICLORAM	10.899	10.801	11.001	702.200	712.500	-1.4

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/12/2024 09/12/2024

Client Sample No.: CCAL03 Date Analyzed: 09/13/2024

Lab Sample No.: HSTDCCC750 Data File : PS027660.D Time Analyzed: 00:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.171	10.070	10.270	687.960	712.500	-3.4
2,4,5-TP(Silvex)	9.755	9.655	9.855	698.960	712.500	-1.9
2,4-D	8.861	8.761	8.961	677.080	705.000	-4.0
2,4-DB	10.734	10.633	10.833	687.720	712.500	-3.5
2,4-DCAA	7.633	7.533	7.733	717.510	750.000	-4.3
3,5-DICHLOROBENZOIC ACID	6.604	6.504	6.704	676.690	697.500	-3.0
4-Nitrophenol	7.167	7.068	7.268	662.840	682.500	-2.9
Dalapon	2.630	2.520	2.720	662.460	682.500	-2.9
DCPA	12.146	12.045	12.245	702.370	720.000	-2.4
DICAMBA	7.828	7.728	7.928	710.100	705.000	0.7
DICHLORPROP	8.535	8.435	8.635	693.830	705.000	-1.6
Dinoseb	11.109	11.009	11.209	679.590	705.000	-3.6
MCPA	8.172	8.072	8.272	67.890	69.750	-2.7
MCPP	7.931	7.831	8.031	72.630	70.500	3.0
Pentachlorophenol	9.378	9.278	9.478	712.570	712.500	0.0
PICLORAM	12.192	12.090	12.290	739.620	712.500	3.8