

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 09/12/2024 09/12/2024

Client Sample No.: CCAL04 Date Analyzed: 09/13/2024

Lab Sample No.: HSTDCCC750 Data File : PS027663.D Time Analyzed: 01:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.341	9.242	9.442	695.170	712.500	-2.4
2,4,5-TP(Silvex)	9.055	8.956	9.156	699.350	712.500	-1.8
2,4-D	8.200	8.101	8.301	683.300	705.000	-3.1
2,4-DB	9.905	9.805	10.005	689.670	712.500	-3.2
2,4-DCAA	7.111	7.012	7.212	723.810	750.000	-3.5
3,5-DICHLOROBENZOIC ACID	6.301	6.201	6.401	678.090	697.500	-2.8
4-Nitrophenol	6.907	6.807	7.007	657.370	682.500	-3.7
Dalapon	2.555	2.454	2.654	654.640	682.500	-4.1
DCPA	11.383	11.283	11.483	707.130	720.000	-1.8
DICAMBA	7.290	7.191	7.391	696.100	705.000	-1.3
DICHLORPROP	7.977	7.878	8.078	679.770	705.000	-3.6
Dinoseb	11.084	10.984	11.184	682.550	705.000	-3.2
MCPA	7.615	7.516	7.716	69.060	69.750	-1.0
MCPP	7.470	7.371	7.571	70.760	70.500	0.4
Pentachlorophenol	8.487	8.388	8.588	695.790	712.500	-2.3
PICLORAM	10.900	10.801	11.001	694.130	712.500	-2.6

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/12/2024 09/12/2024

Client Sample No.: CCAL04 Date Analyzed: 09/13/2024

Lab Sample No.: HSTDCCC750 Data File : PS027663.D Time Analyzed: 01:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	10.171	10.070	10.270	691.300	712.500	-3.0
2,4,5-TP(Silvex)	9.755	9.655	9.855	710.450	712.500	-0.3
2,4-D	8.861	8.761	8.961	682.810	705.000	-3.1
2,4-DB	10.733	10.633	10.833	685.310	712.500	-3.8
2,4-DCAA	7.633	7.533	7.733	716.890	750.000	-4.4
3,5-DICHLOROBENZOIC ACID	6.604	6.504	6.704	685.020	697.500	-1.8
4-Nitrophenol	7.168	7.068	7.268	664.200	682.500	-2.7
Dalapon	2.622	2.520	2.720	657.940	682.500	-3.6
DCPA	12.146	12.045	12.245	720.920	720.000	0.1
DICAMBA	7.828	7.728	7.928	693.030	705.000	-1.7
DICHLORPROP	8.535	8.435	8.635	683.920	705.000	-3.0
Dinoseb	11.110	11.009	11.209	689.480	705.000	-2.2
MCPA	8.172	8.072	8.272	68.170	69.750	-2.3
MCPP	7.931	7.831	8.031	73.380	70.500	4.1
Pentachlorophenol	9.378	9.278	9.478	722.600	712.500	1.4
PICLORAM	12.192	12.090	12.290	724.940	712.500	1.7