

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P3845

Client: Chemtech Consulting Group

Analytical Method: 8151A Datafile : PS027620.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB163250BS	DICAMBA	5	4.90	ug/L	98				67	136	
	MCP	0.5	0.47	ug/L	93				70	130	
	Dalapon	5	4.90	ug/L	98				70	130	
	MCPA	0.5	0.45	ug/L	91				70	130	
	DICHLORPROP	5	4.90	ug/L	98				88	119	
	2,4-D	5	4.90	ug/L	98				83	130	
	2,4,5-TP(Silvex)	5	5.00	ug/L	100				78	127	
	2,4,5-T	5	5.00	ug/L	100				74	129	
	2,4-DB	5	5.00	ug/L	100				53	149	
	Dinoseb	5	5.40	ug/L	108				72	131	
	Pentachlorophenol	5	5.10	ug/L	102				70	130	
	4-Nitrophenol	5	4.80	ug/L	96				70	130	
	PICLORAM	5	4.70	ug/L	94				70	130	
	DCPA	5	5.10	ug/L	102				70	130	
	3,5-DICHLOROBENZOIC	5	4.80	ug/L	96				70	130	