

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ090624

Review By	yogesh	Review On	9/9/2024 8:56:02 AM
Supervise By	Ankita	Supervise On	9/9/2024 9:38:49 AM
SubDirectory	PQ090624	HP Acquire Method	HP Processing Method PQ090624
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,PP23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255		
CCC	PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253		
Internal Standard/PEM			
ICV/I.BLK	PP23259,PP23579,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ068362.D	06 Sep 2024 13:15		YP\AJ	Ok
2	I.BLK	I.BLK	PQ068363.D	06 Sep 2024 13:30		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PQ068364.D	06 Sep 2024 13:44		YP\AJ	Ok
4	AR1660ICC750	AR1660ICC750	PQ068365.D	06 Sep 2024 13:59		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PQ068366.D	06 Sep 2024 14:13		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PQ068367.D	06 Sep 2024 14:28		YP\AJ	Ok
7	AR1660ICC050	AR1660ICC050	PQ068368.D	06 Sep 2024 14:43		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PQ068369.D	06 Sep 2024 14:57		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PQ068370.D	06 Sep 2024 15:12		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PQ068371.D	06 Sep 2024 15:26		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PQ068372.D	06 Sep 2024 15:41		YP\AJ	Ok
12	AR1242ICC500	AR1242ICC500	PQ068373.D	06 Sep 2024 15:55		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PQ068374.D	06 Sep 2024 16:10		YP\AJ	Ok
14	AR1242ICC050	AR1242ICC050	PQ068375.D	06 Sep 2024 16:25		YP\AJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PQ068376.D	06 Sep 2024 16:39		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PQ068377.D	06 Sep 2024 16:54		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PQ068378.D	06 Sep 2024 17:08		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PQ068379.D	06 Sep 2024 17:23		YP\AJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ090624

Review By	yogesh	Review On	9/9/2024 8:56:02 AM
Supervise By	Ankita	Supervise On	9/9/2024 9:38:49 AM
SubDirectory	PQ090624	HP Acquire Method	HP Processing Method PQ090624
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,PP23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255		
CCC	PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253		
Internal Standard/PEM			
ICV/I.BLK	PP23259,PP23579,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PQ068380.D	06 Sep 2024 17:38		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PQ068381.D	06 Sep 2024 17:52		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PQ068382.D	06 Sep 2024 18:07		YPIAJ	Ok
22	AR1254ICC500	AR1254ICC500	PQ068383.D	06 Sep 2024 18:21		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PQ068384.D	06 Sep 2024 18:36		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PQ068385.D	06 Sep 2024 18:51		YPIAJ	Ok
25	AR1262ICC500	AR1262ICC500	PQ068386.D	06 Sep 2024 19:05		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PQ068387.D	06 Sep 2024 19:20		YPIAJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PQ068388.D	06 Sep 2024 19:34		YPIAJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PQ068389.D	06 Sep 2024 19:49		YPIAJ	Ok,M
29	AR1268ICC250	AR1268ICC250	PQ068390.D	06 Sep 2024 20:04		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PQ068391.D	06 Sep 2024 20:18		YPIAJ	Ok,M
31	PQ090624ICV500	ICVPQ090624	PQ068392.D	06 Sep 2024 20:33		YPIAJ	Ok
32	AR1242ICV500	ICVPQ090624AR1242	PQ068393.D	06 Sep 2024 20:47		YPIAJ	Ok,M
33	AR1248ICV500	ICVPQ090624AR1248	PQ068394.D	06 Sep 2024 21:02		YPIAJ	Ok
34	AR1254ICV500	ICVPQ090624AR1254	PQ068395.D	06 Sep 2024 21:17		YPIAJ	Ok
35	AR1268ICV500	ICVPQ090624AR1268	PQ068396.D	06 Sep 2024 21:31		YPIAJ	Ok,M
36	AR1660CCC500	AR1660CCC500	PQ068397.D	06 Sep 2024 21:46		YPIAJ	Ok
37	AR1242CCC500	AR1242CCC500	PQ068398.D	06 Sep 2024 22:00		YPIAJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ090624

Review By	yogesh	Review On	9/9/2024 8:56:02 AM
Supervise By	Ankita	Supervise On	9/9/2024 9:38:49 AM
SubDirectory	PQ090624	HP Acquire Method	HP Processing Method PQ090624
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,PP23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255		
CCC	PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253		
Internal Standard/PEM			
ICV/I.BLK	PP23259,PP23579,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	AR1248CCC500	AR1248CCC500	PQ068399.D	06 Sep 2024 22:15		YPIAJ	Ok
39	AR1254CCC500	AR1254CCC500	PQ068400.D	06 Sep 2024 22:30		YPIAJ	Ok
40	I.BLK	I.BLK	PQ068401.D	06 Sep 2024 22:44		YPIAJ	Ok
41	PB163176BL	PB163176BL	PQ068402.D	06 Sep 2024 22:59		YPIAJ	Ok
42	PB163176BS	PB163176BS	PQ068403.D	06 Sep 2024 23:14		YPIAJ	Ok
43	P3852-01	D3656	PQ068404.D	06 Sep 2024 23:28		YPIAJ	Ok
44	P3852-02	VNJ223	PQ068405.D	06 Sep 2024 23:43		YPIAJ	Ok
45	P3852-02MS	VNJ223MS	PQ068406.D	06 Sep 2024 23:57		YPIAJ	Ok,M
46	P3852-02MSD	VNJ223MSD	PQ068407.D	07 Sep 2024 00:12		YPIAJ	Ok,M
47	P3852-04	ETGI356	PQ068408.D	07 Sep 2024 00:27		YPIAJ	Ok
48	P3857-01	WC-C1-COMP	PQ068409.D	07 Sep 2024 00:41		YPIAJ	Ok
49	P3857-04	WC-C2-COMP	PQ068410.D	07 Sep 2024 00:56		YPIAJ	Ok
50	P3857-07	WC-D1-COMP	PQ068411.D	07 Sep 2024 01:11		YPIAJ	Ok
51	AR1660CCC500	AR1660CCC500	PQ068412.D	07 Sep 2024 02:24		YPIAJ	Ok
52	AR1242CCC500	AR1242CCC500	PQ068413.D	07 Sep 2024 02:38		YPIAJ	Ok
53	AR1248CCC500	AR1248CCC500	PQ068414.D	07 Sep 2024 02:53		YPIAJ	Ok
54	AR1254CCC500	AR1254CCC500	PQ068415.D	07 Sep 2024 03:07		YPIAJ	Ok
55	I.BLK	I.BLK	PQ068416.D	07 Sep 2024 03:22		YPIAJ	Ok
56	P3857-10	WC-D2-COMP	PQ068417.D	07 Sep 2024 03:37		YPIAJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ090624

Review By	yogesh	Review On	9/9/2024 8:56:02 AM
Supervise By	Ankita	Supervise On	9/9/2024 9:38:49 AM
SubDirectory	PQ090624	HP Acquire Method	HP Processing Method PQ090624
STD. NAME	STD REF.#		
Tune/Reschk	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,P P23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP 23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255 PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253 PP23259,PP23579,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

57	P3861-01	DN-B-37	PQ068418.D	07 Sep 2024 03:51		YP\AJ	Ok
58	P3866-01	MR-CAM-04	PQ068419.D	07 Sep 2024 04:06	AR1254 Hit,	YP\AJ	Ok,M
59	P3852-06	D3659	PQ068420.D	07 Sep 2024 04:20		YP\AJ	Ok
60	P3850-01	50428	PQ068421.D	07 Sep 2024 04:35	AR1242 Hit, Need 5x Dilution	YP\AJ	Dilution
61	P3853-01	252805	PQ068422.D	07 Sep 2024 04:50	need cleanup	YP\AJ	Not Ok
62	PB163181BL	PB163181BL	PQ068423.D	07 Sep 2024 05:04		YP\AJ	Ok
63	PB163181BS	PB163181BS	PQ068424.D	07 Sep 2024 05:19		YP\AJ	Ok,M
64	PB163181BSD	PB163181BSD	PQ068425.D	07 Sep 2024 05:34		YP\AJ	Ok
65	AR1660CCC500	AR1660CCC500	PQ068426.D	07 Sep 2024 06:47		YP\AJ	Ok,M
66	AR1242CCC500	AR1242CCC500	PQ068427.D	07 Sep 2024 07:01		YP\AJ	Ok
67	AR1248CCC500	AR1248CCC500	PQ068428.D	07 Sep 2024 07:16		YP\AJ	Ok
68	AR1254CCC500	AR1254CCC500	PQ068429.D	07 Sep 2024 07:30		YP\AJ	Ok
69	I.BLK	I.BLK	PQ068430.D	07 Sep 2024 07:45		YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ090924

Review By	yogesh	Review On	9/10/2024 8:42:34 AM
Supervise By	Ankita	Supervise On	9/10/2024 11:08:22 AM
SubDirectory	PQ090924	HP Acquire Method	HP Processing Method PQ090624

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,PP23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255
CCC	PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253
Internal Standard/PEM	
ICV/I.BLK	PP23259,PP23579,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ068431.D	09 Sep 2024 09:11		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PQ068432.D	09 Sep 2024 09:25		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PQ068433.D	09 Sep 2024 09:56		YP\AJ	Ok
4	AR1248CCC500	AR1248CCC500	PQ068434.D	09 Sep 2024 10:11		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PQ068435.D	09 Sep 2024 10:25		YP\AJ	Ok
6	I.BLK	I.BLK	PQ068436.D	09 Sep 2024 10:40		YP\AJ	Ok
7	P3851-02	BC122669-1-2	PQ068437.D	09 Sep 2024 10:55	AR1254 HIT	YP\AJ	Ok,M
8	P3851-08	BC238813-1-2	PQ068438.D	09 Sep 2024 11:09		YP\AJ	Ok
9	P3851-12	BC271246-1-2	PQ068439.D	09 Sep 2024 11:24		YP\AJ	Ok,M
10	P3851-14	BC274767-1-2	PQ068440.D	09 Sep 2024 11:38		YP\AJ	Ok,M
11	AR1660CCC500	AR1660CCC500	PQ068441.D	09 Sep 2024 17:29		YP\AJ	Ok
12	AR1242CCC500	AR1242CCC500	PQ068442.D	09 Sep 2024 17:43		YP\AJ	Ok
13	AR1248CCC500	AR1248CCC500	PQ068443.D	09 Sep 2024 17:58		YP\AJ	Ok
14	AR1254CCC500	AR1254CCC500	PQ068444.D	09 Sep 2024 18:13		YP\AJ	Ok
15	I.BLK	I.BLK	PQ068445.D	09 Sep 2024 18:27		YP\AJ	Ok
16	PB163161BL	PB163161BL	PQ068446.D	09 Sep 2024 18:42	wrong vial made	YP\AJ	Not Ok
17	P3837-02	RAIN-TANK	PQ068447.D	09 Sep 2024 18:56		YP\AJ	Ok
18	P3837-02RE	RAIN-TANKRE	PQ068448.D	09 Sep 2024 19:11	not required	YP\AJ	Not Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ090924

Review By	yogesh	Review On	9/10/2024 8:42:34 AM
Supervise By	Ankita	Supervise On	9/10/2024 11:08:22 AM
SubDirectory	PQ090924	HP Acquire Method	HP Processing Method PQ090624
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,PP23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255		
CCC	PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253		
Internal Standard/PEM			
ICV/I.BLK	PP23259,PP23579,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PB163181BL	PB163181BL	PQ068449.D	09 Sep 2024 19:26	wrong vial made	YPIAJ	Not Ok
20	P3853-01	252805	PQ068450.D	09 Sep 2024 19:40		YPIAJ	Ok
21	P3850-01DL	50428DL	PQ068451.D	09 Sep 2024 19:55	AR1242 Hit,	YPIAJ	Ok
22	PB163216BL	PB163216BL	PQ068452.D	09 Sep 2024 20:09		YPIAJ	Ok
23	PB163216BS	PB163216BS	PQ068453.D	09 Sep 2024 20:24		YPIAJ	Ok
24	P3887-01	60201	PQ068454.D	09 Sep 2024 20:39	AR1242 Hit, Need 20x Dilution	YPIAJ	Dilution
25	P3887-02	60202	PQ068455.D	09 Sep 2024 20:53	AR1242 Hit, Need 10x Dilution	YPIAJ	Dilution
26	AR1660CCC500	AR1660CCC500	PQ068456.D	09 Sep 2024 22:06		YPIAJ	Ok
27	AR1242CCC500	AR1242CCC500	PQ068457.D	09 Sep 2024 22:21		YPIAJ	Ok
28	AR1248CCC500	AR1248CCC500	PQ068458.D	09 Sep 2024 22:35		YPIAJ	Ok
29	AR1254CCC500	AR1254CCC500	PQ068459.D	09 Sep 2024 22:50		YPIAJ	Ok
30	I.BLK	I.BLK	PQ068460.D	09 Sep 2024 23:04		YPIAJ	Ok
31	P3897-01	1	PQ068461.D	09 Sep 2024 23:19	AR1254 Hit	YPIAJ	Ok,M
32	P3897-02	2	PQ068462.D	09 Sep 2024 23:34	AR1254 Hit	YPIAJ	Ok,M
33	P3897-03	3	PQ068463.D	09 Sep 2024 23:48	AR1254 Hit	YPIAJ	Ok,M
34	P3897-04	4	PQ068464.D	10 Sep 2024 00:03	AR1254 Hit	YPIAJ	Ok,M
35	P3897-05	5	PQ068465.D	10 Sep 2024 00:17	AR1254 Hit	YPIAJ	Ok,M
36	P3897-06	6	PQ068466.D	10 Sep 2024 00:32	AR1254 Hit	YPIAJ	Ok,M
37	P3897-07	7	PQ068467.D	10 Sep 2024 00:47	AR1254 Hit	YPIAJ	Ok,M

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ090924

Review By	yogesh	Review On	9/10/2024 8:42:34 AM
Supervise By	Ankita	Supervise On	9/10/2024 11:08:22 AM
SubDirectory	PQ090924	HP Acquire Method	HP Processing Method PQ090624
STD. NAME	STD REF.#		
Tune/Reschk	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,P P23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP 23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255 PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253 PP23259,PP23579,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	P3897-08	8	PQ068468.D	10 Sep 2024 01:01	AR1254 Hit	YPIAJ	Ok,M
39	PB163217BL	PB163217BL	PQ068469.D	10 Sep 2024 01:16		YPIAJ	Ok
40	PB163217BS	PB163217BS	PQ068470.D	10 Sep 2024 01:30		YPIAJ	Ok
41	AR1660CCC500	AR1660CCC500	PQ068471.D	10 Sep 2024 02:43		YPIAJ	Ok
42	AR1242CCC500	AR1242CCC500	PQ068472.D	10 Sep 2024 02:58		YPIAJ	Ok
43	AR1248CCC500	AR1248CCC500	PQ068473.D	10 Sep 2024 03:13		YPIAJ	Ok
44	AR1254CCC500	AR1254CCC500	PQ068474.D	10 Sep 2024 03:27		YPIAJ	Ok
45	I.BLK	I.BLK	PQ068475.D	10 Sep 2024 03:42		YPIAJ	Ok
46	P3880-01	SHEEPSHEAD-SOIL-P	PQ068476.D	10 Sep 2024 03:56		YPIAJ	Ok
47	P3880-01MS	SHEEPSHEAD-SOIL-P	PQ068477.D	10 Sep 2024 04:11		YPIAJ	Ok,M
48	P3880-01MSD	SHEEPSHEAD-SOIL-P	PQ068478.D	10 Sep 2024 04:25		YPIAJ	Ok,M
49	P3881-01	VNJ-217	PQ068479.D	10 Sep 2024 04:40		YPIAJ	Ok
50	P3888-01	DN-B-36	PQ068480.D	10 Sep 2024 04:55		YPIAJ	Ok,M
51	P3888-02	DN-W-02	PQ068481.D	10 Sep 2024 05:09		YPIAJ	Ok
52	P3892-01	ARS20-005	PQ068482.D	10 Sep 2024 05:24		YPIAJ	Ok
53	P3894-01	SHOP-DRUM	PQ068483.D	10 Sep 2024 05:38	AR1260 Hit	YPIAJ	Ok,M
54	P3896-02	MOO-23-00293	PQ068484.D	10 Sep 2024 05:53	AR1260 Hit	YPIAJ	Ok,M
55	PB163234BL	PB163234BL	PQ068485.D	10 Sep 2024 06:07		YPIAJ	Ok
56	AR1660CCC500	AR1660CCC500	PQ068486.D	10 Sep 2024 07:20		YPIAJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ090924

Review By	yogesh	Review On	9/10/2024 8:42:34 AM
Supervise By	Ankita	Supervise On	9/10/2024 11:08:22 AM
SubDirectory	PQ090924	HP Acquire Method	HP Processing Method PQ090624
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,P P23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP 23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255		
CCC	PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253		
Internal Standard/PEM			
ICV/I.BLK	PP23259,PP23579,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

57	AR1242CCC500	AR1242CCC500	PQ068487.D	10 Sep 2024 07:35		YPIAJ	Ok
58	AR1248CCC500	AR1248CCC500	PQ068488.D	10 Sep 2024 07:49		YPIAJ	Ok
59	AR1254CCC500	AR1254CCC500	PQ068489.D	10 Sep 2024 08:04		YPIAJ	Ok
60	I.BLK	I.BLK	PQ068490.D	10 Sep 2024 08:19		YPIAJ	Ok
61	PB163234BS	PB163234BS	PQ068491.D	10 Sep 2024 08:33		YPIAJ	Ok
62	PB163234BSD	PB163234BSD	PQ068492.D	10 Sep 2024 08:48		YPIAJ	Ok
63	P3845-15	PT-PCBW-WP	PQ068493.D	10 Sep 2024 09:02	AR1221 Hit , AR1221 CCC Missing	YPIAJ	Not Ok
64	PP23641	PP23641	PQ068494.D	10 Sep 2024 09:18		YPIAJ	Ok
65	P3872-02	WC-W-07-202409	PQ068495.D	10 Sep 2024 09:32		YPIAJ	Ok
66	P3896-01	MOO-23-00292	PQ068496.D	10 Sep 2024 09:47		YPIAJ	Ok
67	PP23640	PP23640	PQ068497.D	10 Sep 2024 10:01		YPIAJ	Ok
68	AR1660CCC500	AR1660CCC500	PQ068498.D	10 Sep 2024 10:31		YPIAJ	Ok
69	I.BLK	I.BLK	PQ068499.D	10 Sep 2024 10:45		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ091024

Review By	yogesh	Review On	9/11/2024 8:42:57 AM
Supervise By	Ankita	Supervise On	9/12/2024 8:56:21 AM
SubDirectory	PQ091024	HP Acquire Method	HP Processing Method PQ090624

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,PP23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255
CCC	PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253
Internal Standard/PEM	
ICV/I.BLK	PP23259,PP23579,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PQ068500.D	10 Sep 2024 11:00		YPIAJ	Ok
2	I.BLK	I.BLK	PQ068501.D	10 Sep 2024 11:15		YPIAJ	Ok
3	AR1221ICC1000	AR1221ICC1000	PQ068502.D	10 Sep 2024 11:30		YPIAJ	Ok
4	AR1221ICC750	AR1221ICC750	PQ068503.D	10 Sep 2024 11:44		YPIAJ	Ok
5	AR1221ICC500	AR1221ICC500	PQ068504.D	10 Sep 2024 11:59		YPIAJ	Ok
6	AR1221ICC250	AR1221ICC250	PQ068505.D	10 Sep 2024 12:13		YPIAJ	Ok
7	AR1221ICC050	AR1221ICC050	PQ068506.D	10 Sep 2024 12:28		YPIAJ	Ok
8	AR1221ICV500	ICVPQ091024AR1221	PQ068507.D	10 Sep 2024 12:43		YPIAJ	Ok
9	AR1660CCC500	AR1660CCC500	PQ068508.D	10 Sep 2024 12:57		YPIAJ	Ok
10	AR1221CCC500	AR1221CCC500	PQ068509.D	10 Sep 2024 13:12		YPIAJ	Ok
11	AR1242CCC500	AR1242CCC500	PQ068510.D	10 Sep 2024 13:26		YPIAJ	Ok
12	I.BLK	I.BLK	PQ068511.D	10 Sep 2024 13:41		YPIAJ	Ok
13	P3845-15	PT-PCBW-WP	PQ068512.D	10 Sep 2024 13:56		YPIAJ	Ok
14	P3887-01DL	60201DL	PQ068513.D	10 Sep 2024 14:10	AR1242 Hit,	YPIAJ	Ok
15	P3887-02DL	60202DL	PQ068514.D	10 Sep 2024 14:25	AR1242 Hit,	YPIAJ	Ok
16	PB163253BL	PB163253BL	PQ068515.D	10 Sep 2024 14:39		YPIAJ	Ok
17	PB163253BS	PB163253BS	PQ068516.D	10 Sep 2024 14:54		YPIAJ	Ok
18	PB163253BSD	PB163253BSD	PQ068517.D	10 Sep 2024 15:09		YPIAJ	Ok

Instrument ID: ECD_Q

Daily Analysis Runlog For Sequence/QC Batch ID # PQ091024

Review By	yogesh	Review On	9/11/2024 8:42:57 AM
Supervise By	Ankita	Supervise On	9/12/2024 8:56:21 AM
SubDirectory	PQ091024	HP Acquire Method	HP Processing Method PQ090624
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,PP23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255		
CCC	PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253		
Internal Standard/PEM			
ICV/I.BLK	PP23259,PP23579,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1660CCC500	AR1660CCC500	PQ068518.D	10 Sep 2024 15:23		YPIAJ	Ok
20	AR1221CCC500	AR1221CCC500	PQ068519.D	10 Sep 2024 15:38		YPIAJ	Ok
21	AR1242CCC500	AR1242CCC500	PQ068520.D	10 Sep 2024 15:52		YPIAJ	Ok
22	AR1248CCC500	AR1248CCC500	PQ068521.D	10 Sep 2024 16:07		YPIAJ	Ok
23	AR1254CCC500	AR1254CCC500	PQ068522.D	10 Sep 2024 16:21		YPIAJ	Ok
24	I.BLK	I.BLK	PQ068523.D	10 Sep 2024 16:36		YPIAJ	Ok
25	P3889-01	FMI109	PQ068524.D	10 Sep 2024 16:51	AR1260 Hit	YPIAJ	Ok,M
26	P3912-01	282-COMPOSITE-1	PQ068525.D	10 Sep 2024 17:05		YPIAJ	Ok,M
27	P3912-02	270-COMPOSITE-1	PQ068526.D	10 Sep 2024 17:20		YPIAJ	Ok,M
28	P3913-04	9424-A-C	PQ068527.D	10 Sep 2024 17:35		YPIAJ	Ok
29	P3913-06	9424-D-F	PQ068528.D	10 Sep 2024 17:49	AR1242 Hit , Need 100x Dilution , All Surrogate Fail	YPIAJ	Dilution
30	P3913-08	9424-D-F-82924-D	PQ068529.D	10 Sep 2024 18:04	AR1242 Hit, Need 2x Dilution, DCB Fail in both column	YPIAJ	Dilution
31	P3913-10	82924-A-C	PQ068530.D	10 Sep 2024 18:18	DCB Fail in both column	YPIAJ	ReRun
32	AR1660CCC500	AR1660CCC500	PQ068531.D	10 Sep 2024 19:02		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PQ068532.D	10 Sep 2024 19:17		YPIAJ	Ok
34	AR1248CCC500	AR1248CCC500	PQ068533.D	10 Sep 2024 19:31		YPIAJ	Ok
35	AR1254CCC500	AR1254CCC500	PQ068534.D	10 Sep 2024 19:46		YPIAJ	Ok
36	I.BLK	I.BLK	PQ068535.D	10 Sep 2024 20:00		YPIAJ	Ok

M : Manual Integration