

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

Continuing Calib Date: 09/24/2024 Initial Calibration Date(s): 09/23/2024 09/23/2024

Continuing Calib Time: 17:19 Initial Calibration Time(s): 11:32 12:26

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
alpha-BHC	4.00	4.00	3.90	4.10	0.00
beta-BHC	4.53	4.53	4.43	4.63	0.00
delta-BHC	4.78	4.78	4.68	4.88	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.00
Heptachlor	4.92	4.92	4.82	5.02	0.00
Aldrin	5.26	5.26	5.16	5.36	0.00
Heptachlor epoxide	5.69	5.69	5.59	5.79	0.00
Endosulfan I	6.07	6.08	5.98	6.18	0.01
Dieldrin	6.35	6.35	6.25	6.45	0.00
4,4'-DDE	6.20	6.20	6.10	6.30	0.00
Endrin	6.58	6.58	6.48	6.68	0.00
Endosulfan II	6.80	6.80	6.70	6.90	0.00
4,4'-DDD	6.71	6.71	6.61	6.81	0.00
Endosulfan sulfate	7.16	7.16	7.06	7.26	0.00
4,4'-DDT	7.03	7.03	6.93	7.13	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00
Endrin ketone	7.65	7.65	7.55	7.75	0.00
Endrin aldehyde	6.93	6.93	6.83	7.03	0.00
alpha-Chlordane	6.02	6.02	5.92	6.12	0.00
gamma-Chlordane	5.94	5.95	5.85	6.05	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

Continuing Calib Date: 09/24/2024 Initial Calibration Date(s): 09/23/2024 09/23/2024

Continuing Calib Time: 17:19 Initial Calibration Time(s): 11:32 12:26

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.92	7.92	7.82	8.02	0.00
Tetrachloro-m-xylene	2.78	2.78	2.68	2.88	0.00
alpha-BHC	3.29	3.29	3.19	3.39	0.00
beta-BHC	3.92	3.92	3.82	4.02	0.01
delta-BHC	4.14	4.15	4.05	4.25	0.01
gamma-BHC (Lindane)	3.62	3.62	3.52	3.72	0.00
Heptachlor	3.96	3.96	3.86	4.06	0.01
Aldrin	4.23	4.24	4.14	4.34	0.01
Heptachlor epoxide	4.74	4.74	4.64	4.84	0.00
Endosulfan I	5.11	5.11	5.01	5.21	0.00
Dieldrin	5.37	5.37	5.27	5.47	0.00
4,4'-DDE	5.24	5.24	5.14	5.34	0.00
Endrin	5.65	5.65	5.55	5.75	0.00
Endosulfan II	5.94	5.94	5.84	6.04	0.00
4,4'-DDD	5.80	5.80	5.70	5.90	0.00
Endosulfan sulfate	6.34	6.34	6.24	6.44	0.00
4,4'-DDT	6.05	6.05	5.95	6.15	0.00
Methoxychlor	6.62	6.62	6.52	6.72	0.00
Endrin ketone	6.85	6.85	6.75	6.95	0.00
Endrin aldehyde	6.12	6.12	6.02	6.22	0.00
alpha-Chlordane	5.05	5.05	4.95	5.15	0.00
gamma-Chlordane	4.99	4.99	4.89	5.09	0.00