

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL090924

Review By	Abdul	Review On	9/11/2024 11:44:26 AM
Supervise By	Ankita	Supervise On	9/11/2024 1:52:24 PM
SubDirectory	PL090924	HP Acquire Method	HP Processing Method pl090924 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23282,PP23517
Initial Calibration Stds	PP23473,PP23474,PP23475,PP23476,PP23477,PP23478,PP23479,PP23480,PP23481,PP23482
CCC	PP23484,PP23489,PP23494
Internal Standard/PEM	
ICV/I.BLK	PP23487,PP23497,PP23498
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL091567.D	09 Sep 2024 10:14		AR\AJ	Ok
2	I.BLK	I.BLK	PL091568.D	09 Sep 2024 14:50		AR\AJ	Ok
3	PEM	PEM	PL091569.D	09 Sep 2024 15:03		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL091570.D	09 Sep 2024 15:17		AR\AJ	Ok,M
5	PSTDICC100	PSTDICC100	PL091571.D	09 Sep 2024 15:31		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL091572.D	09 Sep 2024 15:44		AR\AJ	Ok,M
7	PSTDICC050	PSTDICC050	PL091573.D	09 Sep 2024 15:58		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL091574.D	09 Sep 2024 16:11		AR\AJ	Ok,M
9	PSTDICC005	PSTDICC005	PL091575.D	09 Sep 2024 16:25		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL091576.D	09 Sep 2024 16:39		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL091577.D	09 Sep 2024 16:52		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL091578.D	09 Sep 2024 17:06		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL091579.D	09 Sep 2024 17:19		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL091580.D	09 Sep 2024 17:33		AR\AJ	Ok,M
15	PTOXICC1000	PTOXICC1000	PL091581.D	09 Sep 2024 17:47		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL091582.D	09 Sep 2024 18:01		AR\AJ	Ok,M
17	PTOXICC500	PTOXICC500	PL091583.D	09 Sep 2024 18:14		AR\AJ	Ok,M
18	PTOXICC250	PTOXICC250	PL091584.D	09 Sep 2024 18:28		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL090924

Review By	Abdul	Review On	9/11/2024 11:44:26 AM
Supervise By	Ankita	Supervise On	9/11/2024 1:52:24 PM
SubDirectory	PL090924	HP Acquire Method	HP Processing Method pl090924 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23282,PP23517		
Initial Calibration Stds	PP23473,PP23474,PP23475,PP23476,PP23477,PP23478,PP23479,PP23480,PP23481,PP23482		
CCC	PP23484,PP23489,PP23494		
Internal Standard/PEM			
ICV/I.BLK	PP23487,PP23497,PP23498		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL091585.D	09 Sep 2024 18:42		AR\AJ	Ok,M
20	PSTDICV050	ICVPL090924	PL091586.D	09 Sep 2024 18:56		AR\AJ	Ok,M
21	PCHLORICV500	ICVPL090924	PL091587.D	09 Sep 2024 19:10		AR\AJ	Ok
22	PTOXICV500	ICVPL090924	PL091588.D	09 Sep 2024 19:23		AR\AJ	Ok
23	I.BLK	I.BLK	PL091589.D	09 Sep 2024 19:37	F flag coming in second column tcmx, surrogate high in dcb and tcmx in 1st column	AR\AJ	Not Ok
24	PEM	PEM	PL091590.D	09 Sep 2024 19:51		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PL091591.D	09 Sep 2024 20:05		AR\AJ	Ok,M
26	PB163178BS	PB163178BS	PL091592.D	09 Sep 2024 20:18		AR\AJ	Not Ok
27	P3861-01	DN-B-37	PL091593.D	09 Sep 2024 20:32		AR\AJ	Not Ok
28	PB163218BL	PB163218BL	PL091594.D	09 Sep 2024 20:46		AR\AJ	Not Ok
29	PB163218BS	PB163218BS	PL091595.D	09 Sep 2024 21:00		AR\AJ	Not Ok
30	P3888-01	DN-B-36	PL091596.D	09 Sep 2024 21:13	need 2x dilution	AR\AJ	Not Ok
31	P3888-02	DN-W-02	PL091597.D	09 Sep 2024 21:27		AR\AJ	Not Ok
32	P3888-02MS	DN-W-02MS	PL091598.D	09 Sep 2024 21:41		AR\AJ	Not Ok
33	P3888-02MSD	DN-W-02MSD	PL091599.D	09 Sep 2024 21:55		AR\AJ	Not Ok
34	P3892-02	ARS20-005	PL091600.D	09 Sep 2024 22:08	typo P3892-01	AR\AJ	Not Ok
35	I.BLK	I.BLK	PL091601.D	09 Sep 2024 22:22	TCMX HIGH in both column	AR\AJ	Ok,M
36	PSTDCCC050	PSTDCCC050	PL091602.D	09 Sep 2024 22:36		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL091024

Review By	Abdul	Review On	9/11/2024 9:23:53 AM
Supervise By	Ankita	Supervise On	9/12/2024 11:58:50 AM
SubDirectory	PL091024	HP Acquire Method	HP Processing Method p1090924 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23282,PP23517
Initial Calibration Stds	PP23473,PP23474,PP23475,PP23476,PP23477,PP23478,PP23479,PP23480,PP23481,PP23482
CCC	PP23484,PP23489,PP23494
Internal Standard/PEM	
ICV/I.BLK	PP23487,PP23497,PP23498
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL091603.D	10 Sep 2024 10:43		AR\AJ	Ok
2	I.BLK	I.BLK	PL091604.D	10 Sep 2024 10:57		AR\AJ	Ok
3	PEM	PEM	PL091605.D	10 Sep 2024 11:10		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL091606.D	10 Sep 2024 11:24		AR\AJ	Ok,M
5	PCHLORCCC500	PCHLORCCC500	PL091607.D	10 Sep 2024 11:56		AR\AJ	Ok,M
6	PTOXCCC500	PTOXCCC500	PL091608.D	10 Sep 2024 12:12		AR\AJ	Ok,M
7	PB163178BS	PB163178BS	PL091609.D	10 Sep 2024 12:26		AR\AJ	Ok,M
8	PB163229BL	PB163229BL	PL091610.D	10 Sep 2024 12:40		AR\AJ	Ok
9	PB163229BS	PB163229BS	PL091611.D	10 Sep 2024 12:57		AR\AJ	Ok,M
10	PB163229BSD	PB163229BSD	PL091612.D	10 Sep 2024 13:11	Recovery fail	AR\AJ	Not Ok
11	P3845-09	PT-PEST-WP	PL091613.D	10 Sep 2024 13:25	not used	AR\AJ	Not Ok
12	P3853-01	252805	PL091614.D	10 Sep 2024 13:38		AR\AJ	Ok
13	PB163230BL	PB163230BL	PL091615.D	10 Sep 2024 13:52		AR\AJ	Ok
14	PB163230BS	PB163230BS	PL091616.D	10 Sep 2024 15:02	Recovery fail	AR\AJ	Not Ok
15	P3845-11	PT-CHLR-WP	PL091617.D	10 Sep 2024 15:16	not used	AR\AJ	Not Ok
16	PB163231BL	PB163231BL	PL091618.D	10 Sep 2024 15:29		AR\AJ	Ok
17	PB163231BS	PB163231BS	PL091619.D	10 Sep 2024 15:43		AR\AJ	Ok
18	P3845-13	PT-TXP-WP	PL091620.D	10 Sep 2024 15:57	not used	AR\AJ	Not Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL091024

Review By	Abdul	Review On	9/11/2024 9:23:53 AM
Supervise By	Ankita	Supervise On	9/12/2024 11:58:50 AM
SubDirectory	PL091024	HP Acquire Method	HP Processing Method pl090924 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23282,PP23517		
Initial Calibration Stds	PP23473,PP23474,PP23475,PP23476,PP23477,PP23478,PP23479,PP23480,PP23481,PP23482		
CCC	PP23484,PP23489,PP23494		
Internal Standard/PEM			
ICV/I.BLK	PP23487,PP23497,PP23498		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P3845-13DL	PT-TXP-WPDL	PL091621.D	10 Sep 2024 16:23	not used	AR\AJ	Not Ok
20	P3845-13DL2	PT-TXP-WPDL2	PL091622.D	10 Sep 2024 16:43	Not required.	AR\AJ	Not Ok
21	I.BLK	I.BLK	PL091623.D	10 Sep 2024 16:56		AR\AJ	Ok
22	PSTDCCC050	PSTDCCC050	PL091624.D	10 Sep 2024 17:10		AR\AJ	Ok,M
23	PCHLORCCC500	PCHLORCCC500	PL091625.D	10 Sep 2024 17:24		AR\AJ	Ok,M
24	PTOXCCC500	PTOXCCC500	PL091626.D	10 Sep 2024 17:54		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL091124

Review By	Abdul	Review On	9/11/2024 9:41:19 AM				
Supervise By	Ankita	Supervise On	9/12/2024 1:33:13 PM				
SubDirectory	PL091124	HP Acquire Method	HP Processing Method	pl080524 8081			
STD. NAME		STD REF.#					
Tune/Reschk		PP23282,PP23517					
Initial Calibration Stds		PP23473,PP23474,PP23475,PP234876,PP23477,PP23478,PP23479,PP23480,PP23481,PP23482					
CCC		PP23484,PP23489,PP23494					
Internal Standard/PEM							
ICV/I.BLK		PP23487,PP23497,PP23498					
Surrogate Standard							
MS/MSD Standard							
LCS Standard							

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL091627.D	11 Sep 2024 03:04		AR\AJ	Ok
2	I.BLK	I.BLK	PL091628.D	11 Sep 2024 03:18		AR\AJ	Ok
3	PEM	PEM	PL091629.D	11 Sep 2024 03:31		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL091630.D	11 Sep 2024 04:13		AR\AJ	Ok,M
5	P3845-09DL	PT-PEST-WPDL	PL091631.D	11 Sep 2024 04:26	will be reanalyzed	AR\AJ	Not Ok
6	P3861-01	DN-B-37	PL091632.D	11 Sep 2024 04:40		AR\AJ	Ok,M
7	PB163218BL	PB163218BL	PL091633.D	11 Sep 2024 04:54		AR\AJ	Ok
8	PB163218BS	PB163218BS	PL091634.D	11 Sep 2024 05:08	Recovery fail	AR\AJ	Not Ok
9	P3888-01	DN-B-36	PL091635.D	11 Sep 2024 05:21	need 2x dilution	AR\AJ	Dilution
10	P3888-02	DN-W-02	PL091636.D	11 Sep 2024 05:35		AR\AJ	Ok,M
11	P3888-02DL	DN-W-02DL	PL091637.D	11 Sep 2024 05:49	not required	AR\AJ	Not Ok
12	P3888-02MS	DN-W-02MS	PL091638.D	11 Sep 2024 06:02		AR\AJ	Ok,M
13	P3888-02MSD	DN-W-02MSD	PL091639.D	11 Sep 2024 06:16		AR\AJ	Ok,M
14	P3892-01	ARS20-005	PL091640.D	11 Sep 2024 06:30		AR\AJ	Ok,M
15	PB163261BL	PB163261BL	PL091641.D	11 Sep 2024 08:09		AR\AJ	Ok,M
16	PB163261BS	PB163261BS	PL091642.D	11 Sep 2024 09:10		AR\AJ	Ok,M
17	P3872-02	WC-W-07-202409	PL091643.D	11 Sep 2024 09:23		AR\AJ	Ok
18	P3872-02MS	WC-W-07-202409MS	PL091644.D	11 Sep 2024 09:37		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL091124

Review By	Abdul	Review On	9/11/2024 9:41:19 AM
Supervise By	Ankita	Supervise On	9/12/2024 1:33:13 PM
SubDirectory	PL091124	HP Acquire Method	HP Processing Method pl080524 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23282,PP23517		
Initial Calibration Stds	PP23473,PP23474,PP23475,PP234876,PP23477,PP23478,PP23479,PP23480,PP23481,PP23482		
CCC	PP23484,PP23489,PP23494		
Internal Standard/PEM			
ICV/I.BLK	PP23487,PP23497,PP23498		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P3872-02MSD	WC-W-07-202409MSD	PL091645.D	11 Sep 2024 09:51		AR\AJ	Ok,M
20	P3845-09DL	PT-PEST-WPDL	PL091646.D	11 Sep 2024 10:05	TCMX high in both column,F flag coming , Need further dilution	AR\AJ	Not Ok
21	P3905-06	COMP-1	PL091647.D	11 Sep 2024 10:18		AR\AJ	Ok
22	PB163203TB	PB163203TB	PL091648.D	11 Sep 2024 10:32		AR\AJ	Ok
23	PB163205TB	PB163205TB	PL091649.D	11 Sep 2024 10:45		AR\AJ	Ok
24	I.BLK	I.BLK	PL091650.D	11 Sep 2024 10:59		AR\AJ	Ok
25	PSTDCCC050	PSTDCCC050	PL091651.D	11 Sep 2024 11:31		AR\AJ	Ok,M
26	PB163256BL	PB163256BL	PL091652.D	11 Sep 2024 12:05		AR\AJ	Ok
27	PB163256BS	PB163256BS	PL091653.D	11 Sep 2024 12:19		AR\AJ	Ok,M
28	PB163229BSD	PB163229BSD	PL091654.D	11 Sep 2024 13:13		AR\AJ	Ok,M
29	P3906-01	SU-701-COMP-01	PL091655.D	11 Sep 2024 13:27		AR\AJ	Ok,M
30	P3906-07	SU-701-COMP-02	PL091656.D	11 Sep 2024 13:41		AR\AJ	Ok,M
31	P3906-13	SU-701-COMP-03	PL091657.D	11 Sep 2024 13:55	need 10x dilution	AR\AJ	Dilution
32	P3907-01	PL-701-COMP-54	PL091658.D	11 Sep 2024 14:08		AR\AJ	Ok,M
33	P3907-07	PL-701-COMP-55	PL091659.D	11 Sep 2024 14:22		AR\AJ	Ok,M
34	P3916-01	SU-04-090924	PL091660.D	11 Sep 2024 14:36		AR\AJ	Ok,M
35	P3925-01	NB-07091024	PL091661.D	11 Sep 2024 14:50		AR\AJ	Ok,M
36	I.BLK	I.BLK	PL091662.D	11 Sep 2024 15:04		AR\AJ	Ok,M
37	PSTDCCC050	PSTDCCC050	PL091663.D	11 Sep 2024 15:17		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL091124

Review By	Abdul	Review On	9/11/2024 9:41:19 AM
Supervise By	Ankita	Supervise On	9/12/2024 1:33:13 PM
SubDirectory	PL091124	HP Acquire Method	HP Processing Method pl080524 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23282,PP23517		
Initial Calibration Stds	PP23473,PP23474,PP23475,PP234876,PP23477,PP23478,PP23479,PP23480,PP23481,PP23482		
CCC	PP23484,PP23489,PP23494		
Internal Standard/PEM			
ICV/I.BLK	PP23487,PP23497,PP23498		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	PB163290BL	PB163290BL	PL091664.D	11 Sep 2024 15:31	Out of tune time	AR\AJ	Not Ok
39	PB163290BS	PB163290BS	PL091665.D	11 Sep 2024 15:45	Out of tune time	AR\AJ	Not Ok
40	P3917-01	DN-B-35	PL091666.D	11 Sep 2024 15:59	Out of tune time	AR\AJ	Not Ok
41	P3905-01	TP-3A	PL091667.D	11 Sep 2024 16:13	Out of tune time	AR\AJ	Not Ok
42	P3905-05	COMP-1	PL091668.D	11 Sep 2024 16:26	Out of tune time	AR\AJ	Not Ok
43	P3905-05MS	COMP-1MS	PL091669.D	11 Sep 2024 16:40	Out of tune time	AR\AJ	Not Ok
44	P3905-05MSD	COMP-1MSD	PL091670.D	11 Sep 2024 16:54	Out of tune time	AR\AJ	Not Ok
45	I.BLK	I.BLK	PL091671.D	11 Sep 2024 17:08		AR\AJ	Ok
46	PEM	PEM	PL091672.D	11 Sep 2024 17:21		AR\AJ	Ok
47	PSTDCCC050	PSTDCCC050	PL091673.D	11 Sep 2024 17:35		AR\AJ	Ok,M
48	P3929-01	BIN-2	PL091674.D	11 Sep 2024 18:16		AR\AJ	Ok,M
49	P3929-03	72-11930	PL091675.D	11 Sep 2024 18:30		AR\AJ	Ok,M
50	P3929-03MS	72-11930MS	PL091676.D	11 Sep 2024 18:43		AR\AJ	Ok,M
51	P3929-03MSD	72-11930MSD	PL091677.D	11 Sep 2024 18:57		AR\AJ	Ok,M
52	P3930-01	DN-B-33	PL091678.D	11 Sep 2024 19:10		AR\AJ	Ok,M
53	P3930-02	DN-B-34	PL091679.D	11 Sep 2024 19:23		AR\AJ	Ok,M
54	I.BLK	I.BLK	PL091680.D	11 Sep 2024 19:37		AR\AJ	Ok
55	PSTDCCC050	PSTDCCC050	PL091681.D	11 Sep 2024 19:50		AR\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092324

Review By	Abdul	Review On	9/24/2024 9:27:05 AM
Supervise By	Ankita	Supervise On	9/24/2024 11:07:44 AM
SubDirectory	PL092324	HP Acquire Method	HP Processing Method pl092324 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23282,PP23517
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL091952.D	23 Sep 2024 10:38		AR\AJ	Ok
2	I.BLK	I.BLK	PL091953.D	23 Sep 2024 10:52		AR\AJ	Ok
3	PEM	PEM	PL091954.D	23 Sep 2024 11:05		AR\AJ	Ok,M
4	RESCHK	RESCHK	PL091955.D	23 Sep 2024 11:19		AR\AJ	Ok
5	PSTDICC100	PSTDICC100	PL091956.D	23 Sep 2024 11:32		AR\AJ	Ok
6	PSTDICC075	PSTDICC075	PL091957.D	23 Sep 2024 11:45		AR\AJ	Ok
7	PSTDICC050	PSTDICC050	PL091958.D	23 Sep 2024 11:59		AR\AJ	Ok
8	PSTDICC025	PSTDICC025	PL091959.D	23 Sep 2024 12:12		AR\AJ	Ok,M
9	PSTDICC005	PSTDICC005	PL091960.D	23 Sep 2024 12:26		AR\AJ	Ok,M
10	PCHLORICC1000	PCHLORICC1000	PL091961.D	23 Sep 2024 12:39		AR\AJ	Ok
11	PCHLORICC750	PCHLORICC750	PL091962.D	23 Sep 2024 12:52		AR\AJ	Ok
12	PCHLORICC500	PCHLORICC500	PL091963.D	23 Sep 2024 13:06		AR\AJ	Ok
13	PCHLORICC250	PCHLORICC250	PL091964.D	23 Sep 2024 13:19		AR\AJ	Ok
14	PCHLORICC050	PCHLORICC050	PL091965.D	23 Sep 2024 13:33		AR\AJ	Ok
15	PTOXICC1000	PTOXICC1000	PL091966.D	23 Sep 2024 13:46		AR\AJ	Ok
16	PTOXICC750	PTOXICC750	PL091967.D	23 Sep 2024 13:59		AR\AJ	Ok
17	PTOXICC500	PTOXICC500	PL091968.D	23 Sep 2024 14:13		AR\AJ	Ok
18	PTOXICC250	PTOXICC250	PL091969.D	23 Sep 2024 14:26		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092324

Review By	Abdul	Review On	9/24/2024 9:27:05 AM
Supervise By	Ankita	Supervise On	9/24/2024 11:07:44 AM
SubDirectory	PL092324	HP Acquire Method	HP Processing Method
			pl092324 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23282,PP23517		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	PTOXICC100	PTOXICC100	PL091970.D	23 Sep 2024 14:40		AR\AJ	Ok,M
20	PSTDICV050	ICVPL092324	PL091971.D	23 Sep 2024 14:53		AR\AJ	Ok
21	PCHLORICV500	ICVPL092324CHLOR	PL091972.D	23 Sep 2024 15:20		AR\AJ	Ok
22	PTOXICV500	ICVPL092324TOX	PL091973.D	23 Sep 2024 15:47		AR\AJ	Ok
23	I.BLK	I.BLK	PL091974.D	23 Sep 2024 16:13		AR\AJ	Ok
24	PEM	PEM	PL091975.D	23 Sep 2024 16:27		AR\AJ	Ok,M
25	PSTDCCC050	PSTDCCC050	PL091976.D	23 Sep 2024 16:40		AR\AJ	Ok,M
26	PB163572BL	PB163572BL	PL091977.D	23 Sep 2024 16:54		AR\AJ	Ok
27	PB163572BS	PB163572BS	PL091978.D	23 Sep 2024 17:07		AR\AJ	Ok,M
28	P4103-02	WC-22A	PL091979.D	23 Sep 2024 17:20	TCMX low 1st column	AR\AJ	Ok,M
29	P4103-02MS	WC-22AMS	PL091980.D	23 Sep 2024 17:34	TCMX low 1st column	AR\AJ	Ok,M
30	P4103-02MSD	WC-22AMSD	PL091981.D	23 Sep 2024 17:47	TCMX low 1st column	AR\AJ	Ok,M
31	P4103-05	WC-14-5-7.5A	PL091982.D	23 Sep 2024 18:01		AR\AJ	Ok
32	P4103-08	WC-14-5-7.5B	PL091983.D	23 Sep 2024 18:14		AR\AJ	Ok,M
33	P4103-11	WC-14-7.5-10A	PL091984.D	23 Sep 2024 18:28		AR\AJ	Ok,M
34	P4103-14	WC-14-7.5-10B	PL091985.D	23 Sep 2024 18:41		AR\AJ	Ok,M
35	P4103-17	WC-23	PL091986.D	23 Sep 2024 18:55		AR\AJ	Ok,M
36	PB163511TB	PB163511TB	PL091987.D	23 Sep 2024 19:08		AR\AJ	Ok,M
37	I.BLK	I.BLK	PL091988.D	23 Sep 2024 19:21		AR\AJ	Ok,M

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092324

Review By	Abdul	Review On	9/24/2024 9:27:05 AM
Supervise By	Ankita	Supervise On	9/24/2024 11:07:44 AM
SubDirectory	PL092324	HP Acquire Method	HP Processing Method pl092324 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23282,PP23517
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

38	PSTDCCC050	PSTDCCC050	PL091989.D	23 Sep 2024 19:35		ARIAJ	Ok,M
----	------------	------------	------------	-------------------	--	-------	------

M : Manual Integration

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092424

Review By	Abdul	Review On	9/25/2024 10:49:05 AM
Supervise By	Ankita	Supervise On	9/25/2024 2:18:21 PM
SubDirectory	PL092424	HP Acquire Method	HP Processing Method p1092324 8081

STD. NAME	STD REF.#
Tune/Reschk	PP23282,PP23517
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683
CCC	PP23686,PP23690,PP23695
Internal Standard/PEM	
ICV/I.BLK	PP23687,PP23693,PP23698
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PL091990.D	24 Sep 2024 09:15		AR\AJ	Ok
2	I.BLK	I.BLK	PL091991.D	24 Sep 2024 09:28		AR\AJ	Ok
3	PEM	PEM	PL091992.D	24 Sep 2024 09:41		AR\AJ	Ok,M
4	PSTDCCC050	PSTDCCC050	PL091993.D	24 Sep 2024 09:55		AR\AJ	Ok,M
5	PCHLORCCC500	PCHLORCCC500	PL091994.D	24 Sep 2024 12:02		AR\AJ	Ok,M
6	PTOXCCC500	PTOXCCC500	PL091995.D	24 Sep 2024 12:16		AR\AJ	Ok
7	P3845-09	PT-PEST-WP	PL091996.D	24 Sep 2024 12:35	TCMX high in 1st column, Need dilution	AR\AJ	Dilution
8	P3845-11	PT-CHLR-WP	PL091997.D	24 Sep 2024 12:49	TCMX high in 2nd column	AR\AJ	Ok,M
9	P3845-13	PT-TXP-WP	PL091998.D	24 Sep 2024 13:02	Need Dilution	AR\AJ	Dilution
10	P3845-09DL	PT-PEST-WPDL	PL091999.D	24 Sep 2024 13:19	TCMX high in first column , need dilution.	AR\AJ	Dilution
11	P3845-09DL2	PT-PEST-WPDL2	PL092000.D	24 Sep 2024 13:32	TCMX high in first column	AR\AJ	Ok,M
12	P3845-13DL	PT-TXP-WPDL	PL092001.D	24 Sep 2024 13:54		AR\AJ	Ok,M
13	I.BLK	I.BLK	PL092002.D	24 Sep 2024 14:08		AR\AJ	Ok
14	PSTDCCC050	PSTDCCC050	PL092003.D	24 Sep 2024 14:21		AR\AJ	Ok,M
15	PCHLORCCC500	PCHLORCCC500	PL092004.D	24 Sep 2024 14:35		AR\AJ	Ok
16	PTOXCCC500	PTOXCCC500	PL092005.D	24 Sep 2024 14:48		AR\AJ	Ok
17	PB163645BL	PB163645BL	PL092006.D	24 Sep 2024 15:01		AR\AJ	Ok

Instrument ID: ECD_L

Daily Analysis Runlog For Sequence/QC Batch ID # PL092424

Review By	Abdul	Review On	9/25/2024 10:49:05 AM
Supervise By	Ankita	Supervise On	9/25/2024 2:18:21 PM
SubDirectory	PL092424	HP Acquire Method	HP Processing Method pl092324 8081
STD. NAME	STD REF.#		
Tune/Reschk	PP23282,PP23517		
Initial Calibration Stds	PP23673,PP23674,PP23675,PP23676,PP23677,PP23678,PP23679,PP23680,PP23681,PP23682,PP23683		
CCC	PP23686,PP23690,PP23695		
Internal Standard/PEM			
ICV/I.BLK	PP23687,PP23693,PP23698		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	PB163645BS	PB163645BS	PL092007.D	24 Sep 2024 15:32	Some compound Recovery fail	AR\AJ	Not Ok
19	P4149-13	COMP-6	PL092008.D	24 Sep 2024 15:45		AR\AJ	Ok,M
20	P4149-14	COMP-7	PL092009.D	24 Sep 2024 15:59	Need dilution	AR\AJ	Dilution
21	P4149-15	COMP-8	PL092010.D	24 Sep 2024 16:12		AR\AJ	Ok,M
22	P4149-15MS	COMP-8MS	PL092011.D	24 Sep 2024 16:26		AR\AJ	Ok,M
23	P4149-15MSD	COMP-8MSD	PL092012.D	24 Sep 2024 16:39		AR\AJ	Ok,M
24	PB163645BS	PB163645BS	PL092013.D	24 Sep 2024 16:52		AR\AJ	Ok,M
25	I.BLK	I.BLK	PL092014.D	24 Sep 2024 17:06		AR\AJ	Ok
26	PSTDCCC050	PSTDCCC050	PL092015.D	24 Sep 2024 17:19		AR\AJ	Ok,M
27	P4142-01	OR-02-09202024	PL092016.D	24 Sep 2024 05:46 pm		AR\AJ	Ok,M
28	P4146-01	CL-02-092024	PL092017.D	24 Sep 2024 18:00		AR\AJ	Ok,M
29	P4148-01	STOCK-PILE	PL092018.D	24 Sep 2024 18:13		AR\AJ	Ok,M
30	P4150-01	OK-02-09132024	PL092019.D	24 Sep 2024 18:27		AR\AJ	Ok
31	I.BLK	I.BLK	PL092020.D	24 Sep 2024 18:40		AR\AJ	Ok
32	PSTDCCC050	PSTDCCC050	PL092021.D	24 Sep 2024 19:20		AR\AJ	Ok,M

M : Manual Integration