

Surrogate Summary

SDG No.: **P3845**

Client: **Chemtech Consulting Group**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PL091568.D	PIBLK-PL091568.D	Decachlorobiphenyl	1	20	26.7	133		43	140
		Tetrachloro-m-xylene	1	20	24.2	121		77	126
		Decachlorobiphenyl	2	20	26.2	131		43	140
		Tetrachloro-m-xylene	2	20	24.3	121		77	126
I.BLK-PL091604.D	PIBLK-PL091604.D	Decachlorobiphenyl	1	20	18.4	92		43	140
		Tetrachloro-m-xylene	1	20	19.0	95		77	126
		Decachlorobiphenyl	2	20	17.3	87		43	140
		Tetrachloro-m-xylene	2	20	18.5	93		77	126
PB163229BL	PB163229BL	Decachlorobiphenyl	1	20	22.0	110		43	140
		Tetrachloro-m-xylene	1	20	22.5	112		77	126
		Decachlorobiphenyl	2	20	23.3	116		43	140
		Tetrachloro-m-xylene	2	20	21.5	107		77	126
PB163229BS	PB163229BS	Decachlorobiphenyl	1	20	20.9	105		43	140
		Tetrachloro-m-xylene	1	20	19.7	99		77	126
		Decachlorobiphenyl	2	20	21.1	106		43	140
		Tetrachloro-m-xylene	2	20	19.7	99		77	126
I.BLK-PL091623.D	PIBLK-PL091623.D	Decachlorobiphenyl	1	20	20.3	102		43	140
		Tetrachloro-m-xylene	1	20	19.6	98		77	126
		Decachlorobiphenyl	2	20	20.0	100		43	140
		Tetrachloro-m-xylene	2	20	17.9	90		77	126
I.BLK-PL091650.D	PIBLK-PL091650.D	Decachlorobiphenyl	1	20	20.8	104		43	140
		Tetrachloro-m-xylene	1	20	20.2	101		77	126
		Decachlorobiphenyl	2	20	20.4	102		43	140
		Tetrachloro-m-xylene	2	20	19.5	97		77	126
PB163229BSD	PB163229BSD	Decachlorobiphenyl	1	20	21.5	108		43	140
		Tetrachloro-m-xylene	1	20	17.8	89		77	126
		Decachlorobiphenyl	2	20	20.0	100		43	140
		Tetrachloro-m-xylene	2	20	17.5	87		77	126
I.BLK-PL091662.D	PIBLK-PL091662.D	Decachlorobiphenyl	1	20	19.0	95		43	140
		Tetrachloro-m-xylene	1	20	20.1	101		77	126
		Decachlorobiphenyl	2	20	16.4	82		43	140
		Tetrachloro-m-xylene	2	20	19.2	96		77	126
I.BLK-PL091953.D	PIBLK-PL091953.D	Decachlorobiphenyl	1	20	19.6	98		43	140
		Tetrachloro-m-xylene	1	20	19.0	95		77	126
		Decachlorobiphenyl	2	20	18.7	94		43	140
		Tetrachloro-m-xylene	2	20	18.3	92		77	126
I.BLK-PL091991.D	PIBLK-PL091991.D	Decachlorobiphenyl	1	20	20.7	104		43	140
		Tetrachloro-m-xylene	1	20	20.3	101		77	126
		Decachlorobiphenyl	2	20	19.5	98		43	140
		Tetrachloro-m-xylene	2	20	19.6	98		77	126
P3845-09	PT-PEST-WP	Decachlorobiphenyl	1	20	22.1	111		43	140

Surrogate Summary

SDG No.: **P3845**

Client: **Chemtech Consulting Group**

Analytical Method: **8081B**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
P3845-09	PT-PEST-WP	Tetrachloro-m-xylene	1	20	33.8	169	*	77	126
		Decachlorobiphenyl	2	20	21.9	109		43	140
P3845-09DL	PT-PEST-WPDL	Tetrachloro-m-xylene	2	20	20.4	102		77	126
		Decachlorobiphenyl	1	20	23.9	119		43	140
		Tetrachloro-m-xylene	1	20	33.5	168	*	77	126
		Decachlorobiphenyl	2	20	22.4	112		43	140
		Tetrachloro-m-xylene	2	20	22.2	111		77	126
P3845-09DL2	PT-PEST-WPDL2	Decachlorobiphenyl	1	20	23.4	117		43	140
		Tetrachloro-m-xylene	1	20	35.6	178	*	77	126
		Decachlorobiphenyl	2	20	27.4	137		43	140
		Tetrachloro-m-xylene	2	20	20.8	104		77	126
I.BLK-PL092002.D	PIBLK-PL092002.D	Decachlorobiphenyl	1	20	20.9	104		43	140
		Tetrachloro-m-xylene	1	20	20.0	100		77	126
		Decachlorobiphenyl	2	20	20.4	102		43	140
		Tetrachloro-m-xylene	2	20	19.5	97		77	126