

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: **P3845**

Client: **Chemtech Consulting Group**

Analytical Method: **8081B**

Datafile : PL091654.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB163229BSD	alpha-BHC	0.5	0.47	ug/L	94	10			85	130	20
	beta-BHC	0.5	0.46	ug/L	91	11			83	126	20
	delta-BHC	0.5	0.47	ug/L	93	16			69	141	20
	gamma-BHC (Lindane)	0.5	0.47	ug/L	93	12			82	129	20
	Heptachlor	0.5	0.47	ug/L	94	14			79	127	20
	Aldrin	0.5	0.46	ug/L	92	14			79	126	20
	Heptachlor epoxide	0.5	0.46	ug/L	91	9			81	124	20
	Endosulfan I	0.5	0.48	ug/L	96	9			85	122	20
	Dieldrin	0.5	0.48	ug/L	96	12			83	125	20
	4,4'-DDE	0.5	0.48	ug/L	96	16			80	127	20
	Endrin	0.5	0.44	ug/L	88	18			81	128	20
	Endosulfan II	0.5	0.48	ug/L	96	11			82	123	20
	4,4'-DDD	0.5	0.50	ug/L	99	4			77	131	20
	Endosulfan sulfate	0.5	0.50	ug/L	99	9			76	129	20
	4,4'-DDT	0.5	0.45	ug/L	90	20			80	133	20
	Methoxychlor	0.5	0.46	ug/L	92	15			78	108	20
	Endrin ketone	0.5	0.51	ug/L	103	3			80	131	20
	Endrin aldehyde	0.5	0.50	ug/L	100	8			82	127	20
	alpha-Chlordane	0.5	0.49	ug/L	98	9			82	125	20
	gamma-Chlordane	0.5	0.47	ug/L	95	9			82	125	20