

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

Continuing Calib Date: 09/17/2024 Initial Calibration Date(s): 09/09/2024 09/09/2024

Continuing Calib Time: 12:21 Initial Calibration Time(s): 16:39 17:33

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Chlordane-1	(1)	4.71	4.71	4.61	4.81	0.00
Chlordane-2	(2)	5.24	5.24	5.14	5.34	0.00
Chlordane-3	(3)	5.95	5.95	5.85	6.05	0.00
Chlordane-4	(4)	6.03	6.03	5.93	6.13	0.00
Chlordane-5	(5)	6.88	6.88	6.78	6.98	0.00
Decachlorobiphenyl		9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene		3.55	3.54	3.44	3.64	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

Continuing Calib Date: 09/17/2024 Initial Calibration Date(s): 09/09/2024 09/09/2024

Continuing Calib Time: 12:21 Initial Calibration Time(s): 16:39 17:33

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Chlordane-1 (1)	3.78	3.78	3.68	3.88	0.00
Chlordane-2 (2)	4.36	4.36	4.26	4.46	0.00
Chlordane-3 (3)	4.99	4.99	4.89	5.09	0.00
Chlordane-4 (4)	5.05	5.05	4.95	5.15	0.00
Chlordane-5 (5)	5.95	5.95	5.85	6.05	0.00
Decachlorobiphenyl	7.92	7.93	7.83	8.03	0.01
Tetrachloro-m-xylene	2.78	2.78	2.68	2.88	0.00