

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P3845 **SAS No.:** P3845 **SDG NO.:** P3845

Continuing Calib Date: 09/17/2024 **Initial Calibration Date(s):** 09/09/2024 09/09/2024

Continuing Calib Time: 15:24 **Initial Calibration Time(s):** 16:39 17:33

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P3845 **SAS No.:** P3845 **SDG NO.:** P3845

Continuing Calib Date: 09/17/2024 **Initial Calibration Date(s):** 09/09/2024 09/09/2024

Continuing Calib Time: 15:24 **Initial Calibration Time(s):** 16:39 17:33

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.92	7.93	7.83	8.03	0.01
Tetrachloro-m-xylene	2.78	2.78	2.68	2.88	0.00