

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL01 Date Analyzed: 09/10/2024

Lab Sample No.: PCHLORCCC500 Data File : PL091607.D Time Analyzed: 11:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.707	4.606	4.806	461.790	500.000	-7.6
Chlordane-2	5.238	5.136	5.336	474.810	500.000	-5.0
Chlordane-3	5.946	5.845	6.045	449.540	500.000	-10.1
Chlordane-4	6.030	5.927	6.127	473.290	500.000	-5.3
Chlordane-5	6.879	6.776	6.976	433.120	500.000	-13.4
Decachlorobiphenyl	9.061	8.960	9.160	40.970	50.000	-18.1
Tetrachloro-m-xylene	3.545	3.444	3.644	46.930	50.000	-6.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL01 Date Analyzed: 09/10/2024

Lab Sample No.: PCHLORCCC500 Data File : PL091607.D Time Analyzed: 11:56

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	3.781	3.682	3.882	478.360	500.000	-4.3
Chlordane-2	4.359	4.260	4.460	480.710	500.000	-3.9
Chlordane-3	4.990	4.890	5.090	470.290	500.000	-5.9
Chlordane-4	5.053	4.953	5.153	496.270	500.000	-0.7
Chlordane-5	5.949	5.848	6.048	497.570	500.000	-0.5
Decachlorobiphenyl	7.925	7.825	8.025	46.620	50.000	-6.8
Tetrachloro-m-xylene	2.781	2.683	2.883	59.050	50.000	18.1