

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL02 Date Analyzed: 09/10/2024

Lab Sample No.: PCHLORCCC500 Data File : PL091625.D Time Analyzed: 17:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	4.703	4.606	4.806	523.440	500.000	4.7
Chlordane-2	5.233	5.136	5.336	496.640	500.000	-0.7
Chlordane-3	5.941	5.845	6.045	489.490	500.000	-2.1
Chlordane-4	6.025	5.927	6.127	508.950	500.000	1.8
Chlordane-5	6.873	6.776	6.976	515.610	500.000	3.1
Decachlorobiphenyl	9.056	8.960	9.160	47.380	50.000	-5.2
Tetrachloro-m-xylene	3.542	3.444	3.644	48.960	50.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL02 Date Analyzed: 09/10/2024

Lab Sample No.: PCHLORCCC500 Data File : PL091625.D Time Analyzed: 17:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	3.779	3.682	3.882	485.910	500.000	-2.8
Chlordane-2	4.356	4.260	4.460	492.910	500.000	-1.4
Chlordane-3	4.986	4.890	5.090	482.230	500.000	-3.6
Chlordane-4	5.049	4.953	5.153	506.570	500.000	1.3
Chlordane-5	5.945	5.848	6.048	513.360	500.000	2.7
Decachlorobiphenyl	7.921	7.825	8.025	48.350	50.000	-3.3
Tetrachloro-m-xylene	2.779	2.683	2.883	59.110	50.000	18.2