

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL03 Date Analyzed: 09/17/2024

Lab Sample No.: PCHLORCCC500 Data File : PL091780.D Time Analyzed: 12:21

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	4.708	4.606	4.806	492.060	500.000	-1.6
Chlordane-2	5.239	5.136	5.336	446.860	500.000	-10.6
Chlordane-3	5.947	5.845	6.045	460.150	500.000	-8.0
Chlordane-4	6.030	5.927	6.127	485.080	500.000	-3.0
Chlordane-5	6.879	6.776	6.976	494.360	500.000	-1.1
Decachlorobiphenyl	9.064	8.960	9.160	44.470	50.000	-11.1
Tetrachloro-m-xylene	3.547	3.444	3.644	45.460	50.000	-9.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL03 Date Analyzed: 09/17/2024

Lab Sample No.: PCHLORCCC500 Data File : PL091780.D Time Analyzed: 12:21

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	3.780	3.682	3.882	442.640	500.000	-11.5
Chlordane-2	4.358	4.260	4.460	472.820	500.000	-5.4
Chlordane-3	4.989	4.890	5.090	451.210	500.000	-9.8
Chlordane-4	5.051	4.953	5.153	479.710	500.000	-4.1
Chlordane-5	5.947	5.848	6.048	511.550	500.000	2.3
Decachlorobiphenyl	7.924	7.825	8.025	47.220	50.000	-5.6
Tetrachloro-m-xylene	2.780	2.683	2.883	54.380	50.000	8.8