

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL04 Date Analyzed: 09/17/2024

Lab Sample No.: PCHLORCCC500 Data File : PL091786.D Time Analyzed: 15:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.704	4.606	4.806	510.730	500.000	2.1
Chlordane-2	5.234	5.136	5.336	455.150	500.000	-9.0
Chlordane-3	5.942	5.845	6.045	474.700	500.000	-5.1
Chlordane-4	6.025	5.927	6.127	502.430	500.000	0.5
Chlordane-5	6.875	6.776	6.976	518.230	500.000	3.6
Decachlorobiphenyl	9.058	8.960	9.160	46.530	50.000	-6.9
Tetrachloro-m-xylene	3.542	3.444	3.644	48.180	50.000	-3.6

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL04 Date Analyzed: 09/17/2024

Lab Sample No.: PCHLORCCC500 Data File : PL091786.D Time Analyzed: 15:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	3.779	3.682	3.882	468.620	500.000	-6.3
Chlordane-2	4.357	4.260	4.460	492.680	500.000	-1.5
Chlordane-3	4.987	4.890	5.090	465.600	500.000	-6.9
Chlordane-4	5.050	4.953	5.153	499.490	500.000	-0.1
Chlordane-5	5.946	5.848	6.048	524.960	500.000	5.0
Decachlorobiphenyl	7.922	7.825	8.025	48.720	50.000	-2.6
Tetrachloro-m-xylene	2.780	2.683	2.883	57.360	50.000	14.7