

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL05 Date Analyzed: 09/24/2024

Lab Sample No.: PCHLORCCC500 Data File : PL091994.D Time Analyzed: 12:02

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.712	4.606	4.806	420.770	500.000	-15.8
Chlordane-2	5.244	5.136	5.336	433.840	500.000	-13.2
Chlordane-3	5.954	5.846	6.046	422.680	500.000	-15.5
Chlordane-4	6.036	5.928	6.128	422.580	500.000	-15.5
Chlordane-5	6.885	6.777	6.977	430.890	500.000	-13.8
Decachlorobiphenyl	9.070	8.961	9.161	42.700	50.000	-14.6
Tetrachloro-m-xylene	3.552	3.444	3.644	42.730	50.000	-14.5

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL05 Date Analyzed: 09/24/2024

Lab Sample No.: PCHLORCCC500 Data File : PL091994.D Time Analyzed: 12:02

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	3.782	3.681	3.881	433.970	500.000	-13.2
Chlordane-2	4.360	4.258	4.458	436.080	500.000	-12.8
Chlordane-3	4.991	4.889	5.089	439.750	500.000	-12.1
Chlordane-4	5.054	4.951	5.151	439.980	500.000	-12.0
Chlordane-5	5.950	5.847	6.047	439.270	500.000	-12.1
Decachlorobiphenyl	7.926	7.824	8.024	43.140	50.000	-13.7
Tetrachloro-m-xylene	2.783	2.682	2.882	54.100	50.000	8.2