

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL06 Date Analyzed: 09/24/2024

Lab Sample No.: PCHLORCCC500 Data File : PL092004.D Time Analyzed: 14:35

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Chlordane-1	4.706	4.606	4.806	440.860	500.000	-11.8
Chlordane-2	5.235	5.136	5.336	440.530	500.000	-11.9
Chlordane-3	5.946	5.846	6.046	440.740	500.000	-11.9
Chlordane-4	6.027	5.928	6.128	442.260	500.000	-11.5
Chlordane-5	6.876	6.777	6.977	447.250	500.000	-10.6
Decachlorobiphenyl	9.060	8.961	9.161	44.760	50.000	-10.5
Tetrachloro-m-xylene	3.544	3.444	3.644	44.150	50.000	-11.7

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL06 Date Analyzed: 09/24/2024

Lab Sample No.: PCHLORCCC500 Data File : PL092004.D Time Analyzed: 14:35

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Chlordane-1	3.781	3.681	3.881	455.810	500.000	-8.8
Chlordane-2	4.358	4.258	4.458	454.570	500.000	-9.1
Chlordane-3	4.989	4.889	5.089	461.980	500.000	-7.6
Chlordane-4	5.051	4.951	5.151	459.420	500.000	-8.1
Chlordane-5	5.947	5.847	6.047	458.530	500.000	-8.3
Decachlorobiphenyl	7.923	7.824	8.024	45.170	50.000	-9.7
Tetrachloro-m-xylene	2.782	2.682	2.882	56.090	50.000	12.2