

## Report of Analysis

Client:	Chemtech Consulting Group		Date Collected:	09/09/24	
Project:	NJ Waste Water PT		Date Received:	09/09/24	
Client Sample ID:	PIBLK-PL091568.D		SDG No.:	P3845	
Lab Sample ID:	I.BLK-PL091568.D		Matrix:	WATER	
Analytical Method:	SW8081		% Solid:	0	Decanted:
Sample Wt/Vol:	1000	Units: mL	Final Vol:	10000	uL
Soil Aliquot Vol:		uL	Test:	PESTICIDE Group2	
Extraction Type:			Injection Volume :		
GPC Factor :	1.0	PH :			
Prep Method :	3510C				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PL091568.D	1		09/09/24	PL090924

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
<b>TARGETS</b>						
57-74-9	Chlordane	0.082	U	0.082	0.50	ug/L
<b>SURROGATES</b>						
2051-24-3	Decachlorobiphenyl	26.7		43 - 140	133%	SPK: 20
877-09-8	Tetrachloro-m-xylene	24.3		77 - 126	121%	SPK: 20

### Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

\* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit