

Contract:	<u>CHEM02</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>P3845</u>	SAS No.:	<u>P3845</u>	SDG NO.:	<u>P3845</u>
Instrument ID:	<u>ECD_L</u>	Calibration Date(s):		<u>09/09/2024</u>	<u>09/09/2024</u>		
		Calibration Times:		<u>17:47</u>	<u>18:42</u>		

LAB FILE ID:	RT 1000 =	<u>PL091581.D</u>	RT 750 =	<u>PL091582.D</u>
	RT 500 =	PL091583.D	RT 250 =	PL091584.D
			RT 100 =	PL091585.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

Instrument ID: ECD_L Calibration Date(s): 09/09/2024 09/09/2024

Calibration Times: 17:47 18:42

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PL091581.D</u>	RT 750 = <u>PL091582.D</u>
RT 500 = <u>PL091583.D</u>	RT 250 = <u>PL091584.D</u>	RT 100 = <u>PL091585.D</u>

COMPOUND	RT 1000	RT 750	RT 500	RT 250	RT 100	MEAN RT	RT WINDOW	
							FROM	TO
Decachlorobiphenyl	7.92	7.92	7.92	7.92	7.92	7.92	7.82	8.02
Tetrachloro-m-xylene	2.78	2.78	2.78	2.78	2.78	2.78	2.68	2.88
Toxaphene-1 (1)	5.01	5.01	5.01	5.01	5.01	5.01	4.91	5.11
Toxaphene-2 (2)	5.34	5.34	5.34	5.34	5.34	5.34	5.24	5.44
Toxaphene-3 (3)	6.61	6.61	6.61	6.61	6.61	6.61	6.51	6.71
Toxaphene-4 (4)	6.74	6.74	6.74	6.74	6.74	6.74	6.64	6.84
Toxaphene-5 (5)	7.05	7.05	7.05	7.05	7.04	7.05	6.95	7.15