

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

Continuing Calib Date: 09/24/2024 Initial Calibration Date(s): 09/23/2024 09/23/2024

Continuing Calib Time: 12:16 Initial Calibration Time(s): 13:46 14:40

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Toxaphene-1	(1)	6.24	6.24	6.14	6.34	0.00
Toxaphene-2	(2)	6.45	6.45	6.35	6.55	0.00
Toxaphene-3	(3)	7.06	7.06	6.96	7.16	0.00
Toxaphene-4	(4)	7.16	7.15	7.05	7.25	-0.01
Toxaphene-5	(5)	7.94	7.94	7.84	8.04	0.00
Decachlorobiphenyl		9.06	9.06	8.96	9.16	0.00
Tetrachloro-m-xylene		3.55	3.54	3.44	3.64	-0.01

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

Continuing Calib Date: 09/24/2024 Initial Calibration Date(s): 09/23/2024 09/23/2024

Continuing Calib Time: 12:16 Initial Calibration Time(s): 13:46 14:40

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Toxaphene-1	(1)	5.01	5.01	4.91	5.11	0.00
Toxaphene-2	(2)	5.34	5.34	5.24	5.44	0.00
Toxaphene-3	(3)	6.61	6.61	6.51	6.71	0.00
Toxaphene-4	(4)	6.74	6.74	6.64	6.84	0.00
Toxaphene-5	(5)	7.05	7.05	6.95	7.15	0.00
Decachlorobiphenyl		7.92	7.92	7.82	8.02	0.00
Tetrachloro-m-xylene		2.78	2.78	2.68	2.88	0.00