

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL01 Date Analyzed: 09/10/2024

Lab Sample No.: PTOXCCC500 Data File : PL091608.D Time Analyzed: 12:12

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.066	8.954	9.154	45.850	50.000	-8.3
Tetrachloro-m-xylene	3.549	3.442	3.642	49.170	50.000	-1.7
Toxaphene-1	6.245	6.138	6.338	538.680	500.000	7.7
Toxaphene-2	6.451	6.342	6.542	481.410	500.000	-3.7
Toxaphene-3	7.067	6.959	7.159	518.910	500.000	3.8
Toxaphene-4	7.158	7.050	7.250	481.940	500.000	-3.6
Toxaphene-5	7.943	7.834	8.034	488.310	500.000	-2.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL01 Date Analyzed: 09/10/2024

Lab Sample No.: PTOXCCC500 Data File : PL091608.D Time Analyzed: 12:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.927	7.821	8.021	50.560	50.000	1.1
Tetrachloro-m-xylene	2.783	2.682	2.882	50.060	50.000	0.1
Toxaphene-1	5.014	4.910	5.110	450.730	500.000	-9.9
Toxaphene-2	5.340	5.236	5.436	445.750	500.000	-10.9
Toxaphene-3	6.614	6.509	6.709	515.020	500.000	3.0
Toxaphene-4	6.743	6.638	6.838	545.360	500.000	9.1
Toxaphene-5	7.055	6.948	7.148	453.470	500.000	-9.3