

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL02 Date Analyzed: 09/10/2024

Lab Sample No.: PTOXCCC500 Data File : PL091626.D Time Analyzed: 17:54

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.057	8.954	9.154	49.360	50.000	-1.3
Tetrachloro-m-xylene	3.543	3.442	3.642	51.020	50.000	2.0
Toxaphene-1	6.239	6.138	6.338	561.870	500.000	12.4
Toxaphene-2	6.444	6.342	6.542	453.580	500.000	-9.3
Toxaphene-3	7.060	6.959	7.159	527.520	500.000	5.5
Toxaphene-4	7.152	7.050	7.250	511.460	500.000	2.3
Toxaphene-5	7.936	7.834	8.034	526.200	500.000	5.2

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/09/2024 09/09/2024

Client Sample No.: CCAL02 Date Analyzed: 09/10/2024

Lab Sample No.: PTOXCCC500 Data File : PL091626.D Time Analyzed: 17:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.921	7.821	8.021	50.660	50.000	1.3
Tetrachloro-m-xylene	2.780	2.682	2.882	48.970	50.000	-2.1
Toxaphene-1	5.011	4.910	5.110	452.230	500.000	-9.6
Toxaphene-2	5.336	5.236	5.436	576.710	500.000	15.3
Toxaphene-3	6.609	6.509	6.709	540.390	500.000	8.1
Toxaphene-4	6.738	6.638	6.838	550.730	500.000	10.1
Toxaphene-5	7.050	6.948	7.148	477.520	500.000	-4.5