

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL03 Date Analyzed: 09/24/2024

Lab Sample No.: PTOXCCC500 Data File : PL091995.D Time Analyzed: 12:16

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.061	8.961	9.161	48.700	50.000	-2.6
Tetrachloro-m-xylene	3.545	3.444	3.644	48.220	50.000	-3.6
Toxaphene-1	6.243	6.142	6.342	486.800	500.000	-2.6
Toxaphene-2	6.447	6.346	6.546	479.130	500.000	-4.2
Toxaphene-3	7.064	6.964	7.164	476.210	500.000	-4.8
Toxaphene-4	7.155	7.054	7.254	478.010	500.000	-4.4
Toxaphene-5	7.939	7.839	8.039	482.930	500.000	-3.4

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL03 Date Analyzed: 09/24/2024

Lab Sample No.: PTOXCCC500 Data File : PL091995.D Time Analyzed: 12:16

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.923	7.824	8.024	48.900	50.000	-2.2
Tetrachloro-m-xylene	2.782	2.683	2.883	48.910	50.000	-2.2
Toxaphene-1	5.012	4.913	5.113	500.090	500.000	0.0
Toxaphene-2	5.338	5.238	5.438	507.990	500.000	1.6
Toxaphene-3	6.611	6.512	6.712	500.740	500.000	0.1
Toxaphene-4	6.740	6.640	6.840	477.910	500.000	-4.4
Toxaphene-5	7.051	6.953	7.153	493.250	500.000	-1.4