

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL04 Date Analyzed: 09/24/2024

Lab Sample No.: PTOXCCC500 Data File : PL092005.D Time Analyzed: 14:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.059	8.961	9.161	49.410	50.000	-1.2
Tetrachloro-m-xylene	3.543	3.444	3.644	48.260	50.000	-3.5
Toxaphene-1	6.241	6.142	6.342	497.690	500.000	-0.5
Toxaphene-2	6.445	6.346	6.546	453.660	500.000	-9.3
Toxaphene-3	7.062	6.964	7.164	475.600	500.000	-4.9
Toxaphene-4	7.153	7.054	7.254	478.240	500.000	-4.4
Toxaphene-5	7.938	7.839	8.039	486.760	500.000	-2.6

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/23/2024 09/23/2024

Client Sample No.: CCAL04 Date Analyzed: 09/24/2024

Lab Sample No.: PTOXCCC500 Data File : PL092005.D Time Analyzed: 14:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.922	7.824	8.024	49.160	50.000	-1.7
Tetrachloro-m-xylene	2.782	2.683	2.883	48.980	50.000	-2.0
Toxaphene-1	5.012	4.913	5.113	524.570	500.000	4.9
Toxaphene-2	5.337	5.238	5.438	526.790	500.000	5.4
Toxaphene-3	6.611	6.512	6.712	501.780	500.000	0.4
Toxaphene-4	6.738	6.640	6.840	485.120	500.000	-3.0
Toxaphene-5	7.051	6.953	7.153	484.370	500.000	-3.1