

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

Instrument ID: ECD_R Calibration Date(s): 09/11/2024 09/11/2024
Calibration Times: 11:54 12:44

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 0.5 = <u>PR068520.D</u>		CF 0.25 = <u>PR068521.D</u>			
CF 0.1 = <u>PR068522.D</u>		CF 0.05 = <u>PR068523.D</u>		CF 0.025 = <u>PR068524.D</u>			
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD
DBCP	25120500000	25711200000	26894000000	28515800000	27298600000	26708000000	5
EDB	13761800000	14300500000	15128900000	15480900000	14083200000	14551100000	5

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

Instrument ID: ECD_R Calibration Date(s): 09/11/2024 09/11/2024
Calibration Times: 11:54 12:44

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:		CF 0.5 = <u>PR068520.D</u>		CF 0.25 = <u>PR068521.D</u>			
CF 0.1 = <u>PR068522.D</u>		CF 0.05 = <u>PR068523.D</u>		CF 0.025 = <u>PR068524.D</u>			
COMPOUND	CF 0.5	CF 0.25	CF 0.1	CF 0.05	CF 0.025	CF	% RSD
DBCP	130710000000	133056000000	141053000000	149210000000	137898000000	138385000000	5
EDB	70504900000	72410600000	78010100000	82568200000	77180500000	76134900000	6