

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P3845 **SAS No.:** P3845 **SDG NO.:** P3845

Continuing Calib Date: 09/11/2024 **Initial Calibration Date(s):** 09/11/2024 09/11/2024

Continuing Calib Time: 13:12 **Initial Calibration Time(s):** 11:54 12:44

GC Column: ZB-MR1 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	6.58	6.58	6.48	6.68	0.00
EDB	2.84	2.84	2.74	2.94	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P3845 **SAS No.:** P3845 **SDG NO.:** P3845

Continuing Calib Date: 09/11/2024 **Initial Calibration Date(s):** 09/11/2024 09/11/2024

Continuing Calib Time: 13:12 **Initial Calibration Time(s):** 11:54 12:44

GC Column: ZB-MR2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	5.55	5.56	5.46	5.66	0.01
EDB	2.23	2.23	2.13	2.33	0.00