

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** CHEM02

**Lab Code:** CHEM      **Case No.:** P3845      **SAS No.:** P3845      **SDG NO.:** P3845

**Continuing Calib Date:** 09/11/2024      **Initial Calibration Date(s):** 09/11/2024      09/11/2024

**Continuing Calib Time:** 14:34      **Initial Calibration Time(s):** 11:54      12:44

**GC Column:** ZB-MR1      **ID:** 0.32      (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	6.58	6.58	6.48	6.68	0.00
EDB	2.84	2.84	2.74	2.94	0.00

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** CHEM02

**Lab Code:** CHEM      **Case No.:** P3845      **SAS No.:** P3845      **SDG NO.:** P3845

**Continuing Calib Date:** 09/11/2024      **Initial Calibration Date(s):** 09/11/2024      09/11/2024

**Continuing Calib Time:** 14:34      **Initial Calibration Time(s):** 11:54      12:44

**GC Column:** ZB-MR2      **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
DBCP	5.55	5.56	5.46	5.66	0.01
EDB	2.23	2.23	2.13	2.33	0.00