

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 09/11/2024 09/11/2024

Client Sample No.: CCAL01 Date Analyzed: 09/11/2024

Lab Sample No.: M8011-504 CCC 0. Data File : PR068527.D Time Analyzed: 13:12

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
DBCP	6.578	6.479	6.679	0.100	0.100	0.0
EDB	2.840	2.740	2.940	0.110	0.100	-0.10

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P3845 **SAS No.:** P3845 **SDG NO.:** P3845

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/11/2024 09/11/2024

Client Sample No.: CCAL01 **Date Analyzed:** 09/11/2024

Lab Sample No.: M8011-504 CCC 0. **Data File :** PR068527.D **Time Analyzed:** 13:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	5.554	5.455	5.655	0.100	0.100	0.0
EDB	2.231	2.131	2.331	0.100	0.100	0.0