

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P3845 **SAS No.:** P3845 **SDG NO.:** P3845

GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/11/2024 09/11/2024

Client Sample No.: CCAL02 **Date Analyzed:** 09/11/2024

Lab Sample No.: M8011-504 CCC 0. **Data File :** PR068534.D **Time Analyzed:** 14:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	6.578	6.479	6.679	0.100	0.100	0.0
EDB	2.839	2.740	2.940	0.110	0.100	-0.10

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM **Case No.:** P3845 **SAS No.:** P3845 **SDG NO.:** P3845

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/11/2024 09/11/2024

Client Sample No.: CCAL02 **Date Analyzed:** 09/11/2024

Lab Sample No.: M8011-504 CCC 0. **Data File :** PR068534.D **Time Analyzed:** 14:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
DBCP	5.554	5.455	5.655	0.100	0.100	0.0
EDB	2.230	2.131	2.331	0.100	0.100	0.0