

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091624\
 Data File : BM047533.D
 Acq On : 16 Sep 2024 10:16
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 16 11:33:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM091424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Sep 15 08:56:37 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	108	0.00
2	1,4-Dioxane	40.000	40.292	-0.7	112	0.00
3	Pyridine	40.000	40.479	-1.2	110	0.00
4	n-Nitrosodimethylamine	40.000	39.718	0.7	106	0.00
5 S	2-Fluorophenol	80.000	79.754	0.3	108	0.00
6	Aniline	40.000	40.518	-1.3	108	0.00
7 S	Phenol-d6	80.000	81.134	-1.4	107	0.00
8	2-Chlorophenol	40.000	39.781	0.5	106	0.00
9	Benzaldehyde	40.000	37.547	6.1	101	0.00
10 C	Phenol	40.000	40.304	-0.8	108	0.00
11	bis(2-Chloroethyl)ether	40.000	40.460	-1.2	109	0.00
12	1,3-Dichlorobenzene	40.000	39.313	1.7	107	0.00
13 C	1,4-Dichlorobenzene	40.000	39.231	1.9	107	0.00
14	1,2-Dichlorobenzene	40.000	39.346	1.6	107	0.00
15	Benzyl Alcohol	40.000	40.082	-0.2	105	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.838	0.4	107	0.00
17	2-Methylphenol	40.000	40.301	-0.8	106	0.00
18	Hexachloroethane	40.000	38.895	2.8	105	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	41.470	-3.7	107	0.00
20	3+4-Methylphenols	40.000	39.944	0.1	105	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	106	0.00
22	Acetophenone	40.000	39.840	0.4	105	0.00
23 S	Nitrobenzene-d5	80.000	81.114	-1.4	106	0.00
24	Nitrobenzene	40.000	40.197	-0.5	106	0.00
25	Isophorone	40.000	40.635	-1.6	106	0.00
26 C	2-Nitrophenol	40.000	39.105	2.2	99	0.00
27	2,4-Dimethylphenol	40.000	39.346	1.6	105	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.374	-0.9	107	0.00
29 C	2,4-Dichlorophenol	40.000	40.351	-0.9	105	0.00
30	1,2,4-Trichlorobenzene	40.000	39.325	1.7	106	0.00
31	Naphthalene	40.000	39.680	0.8	105	0.00
32	Benzoic acid	40.000	37.431	6.4	103	0.00
33	4-Chloroaniline	40.000	40.562	-1.4	107	0.00
34 C	Hexachlorobutadiene	40.000	38.483	3.8	103	0.00
35	Caprolactam	40.000	43.526	-8.8	110	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.231	-3.1	106	0.00
37	2-Methylnaphthalene	40.000	40.201	-0.5	105	0.00
38	1-Methylnaphthalene	40.000	40.320	-0.8	107	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	107	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.334	1.7	104	0.00
41 P	Hexachlorocyclopentadiene	40.000	38.172	4.6	99	0.00
42 S	2,4,6-Tribromophenol	80.000	81.414	-1.8	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.025	-0.1	103	0.00
44	2,4,5-Trichlorophenol	40.000	40.261	-0.7	104	0.00
45 S	2-Fluorobiphenyl	80.000	81.679	-2.1	104	0.00
46	1,1'-Biphenyl	40.000	39.946	0.1	106	0.00
47	2-Chloronaphthalene	40.000	39.790	0.5	105	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091624\
 Data File : BM047533.D
 Acq On : 16 Sep 2024 10:16
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 16 11:33:47 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM091424.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Sep 15 08:56:37 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.494	-3.7	105	0.00
49	Acenaphthylene	40.000	40.604	-1.5	106	0.00
50	Dimethylphthalate	40.000	39.962	0.1	106	0.00
51	2,6-Dinitrotoluene	40.000	41.552	-3.9	107	0.00
52 C	Acenaphthene	40.000	40.594	-1.5	106	0.00
53	3-Nitroaniline	40.000	42.399	-6.0	108	0.00
54 P	2,4-Dinitrophenol	40.000	36.995	7.5	95	0.00
55	Dibenzofuran	40.000	40.086	-0.2	106	0.00
56 P	4-Nitrophenol	40.000	42.277	-5.7	105	0.00
57	2,4-Dinitrotoluene	40.000	42.670	-6.7	107	0.00
58	Fluorene	40.000	41.789	-4.5	106	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.362	-0.9	105	0.00
60	Diethylphthalate	40.000	41.046	-2.6	107	0.00
61	4-Chlorophenyl-phenylether	40.000	41.102	-2.8	104	0.00
62	4-Nitroaniline	40.000	44.755	-11.9	107	0.00
63	Azobenzene	40.000	41.753	-4.4	108	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	108	0.00
65	4,6-Dinitro-2-methylphenol	40.000	35.256	11.9	100	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.470	-1.2	106	0.00
67	4-Bromophenyl-phenylether	40.000	40.024	-0.1	107	0.00
68	Hexachlorobenzene	40.000	39.328	1.7	105	0.00
69	Atrazine	40.000	40.736	-1.8	105	0.00
70 C	Pentachlorophenol	40.000	42.539	-6.3	108	0.00
71	Phenanthrene	40.000	40.604	-1.5	107	0.00
72	Anthracene	40.000	41.313	-3.3	107	0.00
73	Carbazole	40.000	41.705	-4.3	109	0.00
74	Di-n-butylphthalate	40.000	42.846	-7.1	109	0.00
75 C	Fluoranthene	40.000	42.745	-6.9	108	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	109	0.00
77	Benzydine	40.000	58.267	-45.7#	176	0.00
78	Pyrene	40.000	40.890	-2.2	109	0.00
79 S	Terphenyl-d14	80.000	89.012	-11.3	108	0.00
80	Butylbenzylphthalate	40.000	40.521	-1.3	105	0.00
81	Benzo(a)anthracene	40.000	40.654	-1.6	108	0.00
82	3,3'-Dichlorobenzidine	40.000	44.650	-11.6	117	0.00
83	Chrysene	40.000	40.493	-1.2	109	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.311	-5.8	109	0.00
85 c	Di-n-octyl phthalate	40.000	42.252	-5.6	110	0.00
86 I	Perylene-d12	20.000	20.000	0.0	114	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.637	0.9	114	0.00
88	Benzo(b)fluoranthene	40.000	42.001	-5.0	116	0.00
89	Benzo(k)fluoranthene	40.000	41.296	-3.2	111	0.00
90 C	Benzo(a)pyrene	40.000	41.189	-3.0	113	0.00
91	Dibenzo(a,h)anthracene	40.000	39.299	1.8	114	0.00
92	Benzo(g,h,i)perylene	40.000	39.113	2.2	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM091624\
Data File : BM047533.D
Acq On : 16 Sep 2024 10:16
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Sep 16 11:33:47 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM091424.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Sep 15 08:56:37 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0