

Surrogate Summary

SW-846

SDG No.: P3845

Client: Chemtech Consulting Group

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P3845-03	PT-BN-WP	2-Fluorophenol	150	153	102		10	139
		Phenol-d6	150	145	97		10	134
		Nitrobenzene-d5	100	96.7	97		49	133
		2-Fluorobiphenyl	100	98.0	98		52	132
		2,4,6-Tribromophenol	150	111	74		44	137
		Terphenyl-d14	100	132	132	*	48	125
P3845-03DL	PT-BN-WPDL	2-Fluorophenol	150	155	103		10	139
		Phenol-d6	150	148	99		10	134
		Nitrobenzene-d5	100	103	103		49	133
		2-Fluorobiphenyl	100	94.7	95		52	132
		2,4,6-Tribromophenol	150	98.8	66		44	137
		Terphenyl-d14	100	119	119		48	125
PB163318BL	PB163318BL	2-Fluorophenol	150	124	83		10	139
		Phenol-d6	150	109	73		10	134
		Nitrobenzene-d5	100	78.0	78		49	133
		2-Fluorobiphenyl	100	76.7	77		52	132
		2,4,6-Tribromophenol	150	107	71		44	137
		Terphenyl-d14	100	95.1	95		48	125
PB163318BS	PB163318BS	2-Fluorophenol	150	156	104		10	139
		Phenol-d6	150	140	93		10	134
		Nitrobenzene-d5	100	96.1	96		49	133
		2-Fluorobiphenyl	100	95.8	96		52	132
		2,4,6-Tribromophenol	150	142	95		44	137
		Terphenyl-d14	100	111	111		48	125
PB163318BSD	PB163318BSD	2-Fluorophenol	150	153	102		10	139
		Phenol-d6	150	139	92		10	134
		Nitrobenzene-d5	100	96.5	96		49	133
		2-Fluorobiphenyl	100	96.9	97		52	132
		2,4,6-Tribromophenol	150	142	95		44	137
		Terphenyl-d14	100	113	113		48	125