

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

EPA Sample No.: SSTDCCC040 Date Analyzed: 09/16/2024

Lab File ID: BM047533.D Time Analyzed: 10:16

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	392657	7.851	1713350	10.65	1011170	14.48
UPPER LIMIT	785314	8.351	3426700	11.145	2022340	14.98
LOWER LIMIT	196329	7.351	856675	10.145	505585	13.98
EPA SAMPLE NO.						
01 PT-BN-WP	344273	7.85	1460340	10.65	837356	14.48
02 PT-BN-WPDL	377823	7.85	1623270	10.65	949493	14.48

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.