

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

EPA Sample No.: SSTDCCC040 Date Analyzed: 09/19/2024

Lab File ID: BM047566.D Time Analyzed: 11:49

Instrument ID: BNA\_M GC Column: ZB-GR ID: 0.25 (mm)

|                | IS1 (DCB)<br>AREA # | RT #  | IS2 (NPT)<br>AREA # | RT #   | IS3 (ANT)<br>AREA # | RT #   |
|----------------|---------------------|-------|---------------------|--------|---------------------|--------|
| 12 HOUR STD    | 406101              | 7.839 | 1674500             | 10.64  | 905170              | 14.47  |
| UPPER LIMIT    | 812202              | 8.339 | 3349000             | 11.139 | 1810340             | 14.974 |
| LOWER LIMIT    | 203051              | 7.339 | 837250              | 10.139 | 452585              | 13.974 |
| EPA SAMPLE NO. |                     |       |                     |        |                     |        |
| 01 PB163318BL  | 353821              | 7.84  | 1340230             | 10.63  | 659710              | 14.47  |
| 02 PB163318BS  | 304822              | 7.84  | 1201950             | 10.64  | 615193              | 14.47  |
| 03 PB163318BSD | 312652              | 7.85  | 1211920             | 10.64  | 613404              | 14.47  |

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.

\* Values outside of QC limits.