

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG NO.: P3845

EPA Sample No.: SSTDCCC040 Date Analyzed: 09/16/2024

Lab File ID: BM047533.D Time Analyzed: 10:16

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	1938190	17.215	1774640	21.439	1492870	24.468
UPPER LIMIT	3876380	17.715	3549280	21.939	2985740	24.968
LOWER LIMIT	969095	16.715	887320	20.939	746435	23.968
EPA SAMPLE NO.						
01 PT-BN-WP	1576550	17.22	1292100	21.44	1275490	24.47
02 PT-BN-WPDL	1818860	17.22	1347120	21.44	1275890	24.47

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.