

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091624\
 Data File : BF139377.D
 Acq On : 16 Sep 2024 10:36
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 16 11:10:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 14 02:58:23 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	125	0.00
2	1,4-Dioxane	40.000	40.309	-0.8	126	0.00
3	Pyridine	40.000	42.424	-6.1	131	0.00
4	n-Nitrosodimethylamine	40.000	39.320	1.7	123	0.00
5 S	2-Fluorophenol	80.000	79.476	0.7	126	0.00
6	Aniline	40.000	39.081	2.3	125	-0.02
7 S	Phenol-d6	80.000	78.245	2.2	123	0.00
8	2-Chlorophenol	40.000	39.525	1.2	126	0.00
9	Benzaldehyde	40.000	34.698	13.3	121	0.00
10 C	Phenol	40.000	39.765	0.6	127	0.00
11	bis(2-Chloroethyl)ether	40.000	37.579	6.1	129	0.00
12	1,3-Dichlorobenzene	40.000	39.401	1.5	125	0.00
13 C	1,4-Dichlorobenzene	40.000	39.787	0.5	126	0.00
14	1,2-Dichlorobenzene	40.000	39.769	0.6	127	0.00
15	Benzyl Alcohol	40.000	39.435	1.4	124	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.896	-2.2	132	0.00
17	2-Methylphenol	40.000	39.729	0.7	127	0.00
18	Hexachloroethane	40.000	39.803	0.5	127	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.524	-1.3	129	0.00
20	3+4-Methylphenols	40.000	39.204	2.0	124	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	127	0.00
22	Acetophenone	40.000	41.115	-2.8	126	0.00
23 S	Nitrobenzene-d5	80.000	83.307	-4.1	127	0.00
24	Nitrobenzene	40.000	40.917	-2.3	125	0.00
25	Isophorone	40.000	41.226	-3.1	128	0.00
26 C	2-Nitrophenol	40.000	41.490	-3.7	123	0.00
27	2,4-Dimethylphenol	40.000	41.426	-3.6	125	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.886	-2.2	128	0.00
29 C	2,4-Dichlorophenol	40.000	40.675	-1.7	125	0.00
30	1,2,4-Trichlorobenzene	40.000	40.040	-0.1	124	0.00
31	Naphthalene	40.000	40.811	-2.0	126	0.00
32	Benzoic acid	40.000	41.552	-3.9	126	0.00
33	4-Chloroaniline	40.000	40.580	-1.4	122	0.00
34 C	Hexachlorobutadiene	40.000	41.645	-4.1	129	0.00
35	Caprolactam	40.000	39.424	1.4	117	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.528	-1.3	123	0.00
37	2-Methylnaphthalene	40.000	40.440	-1.1	125	0.00
38	1-Methylnaphthalene	40.000	40.060	-0.2	125	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	124	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	41.894	-4.7	124	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.327	-10.8	124	0.00
42 S	2,4,6-Tribromophenol	80.000	78.227	2.2	113	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.099	-5.2	119	0.00
44	2,4,5-Trichlorophenol	40.000	41.068	-2.7	119	0.00
45 S	2-Fluorobiphenyl	80.000	83.451	-4.3	124	0.00
46	1,1'-Biphenyl	40.000	41.446	-3.6	123	0.00
47	2-Chloronaphthalene	40.000	41.408	-3.5	122	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091624\
 Data File : BF139377.D
 Acq On : 16 Sep 2024 10:36
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Sep 16 11:10:45 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Sep 14 02:58:23 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.713	-4.3	121	0.00
49	Acenaphthylene	40.000	40.863	-2.2	121	0.00
50	Dimethylphthalate	40.000	40.713	-1.8	120	0.00
51	2,6-Dinitrotoluene	40.000	42.326	-5.8	122	0.00
52 C	Acenaphthene	40.000	40.866	-2.2	119	0.00
53	3-Nitroaniline	40.000	43.067	-7.7	122	0.00
54 P	2,4-Dinitrophenol	40.000	38.724	3.2	109	0.00
55	Dibenzofuran	40.000	40.236	-0.6	118	0.00
56 P	4-Nitrophenol	40.000	40.098	-0.2	113	0.00
57	2,4-Dinitrotoluene	40.000	41.104	-2.8	115	0.00
58	Fluorene	40.000	40.449	-1.1	117	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.442	-3.6	120	0.00
60	Diethylphthalate	40.000	40.532	-1.3	118	0.00
61	4-Chlorophenyl-phenylether	40.000	40.879	-2.2	120	0.00
62	4-Nitroaniline	40.000	39.234	1.9	111	0.00
63	Azobenzene	40.000	40.634	-1.6	118	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	114	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.085	-7.7	111	0.00
66 c	n-Nitrosodiphenylamine	40.000	41.895	-4.7	116	0.00
67	4-Bromophenyl-phenylether	40.000	43.182	-8.0	118	0.00
68	Hexachlorobenzene	40.000	41.886	-4.7	114	0.00
69	Atrazine	40.000	34.691	13.3	113	0.00
70 C	Pentachlorophenol	40.000	43.037	-7.6	114	0.00
71	Phenanthrene	40.000	41.230	-3.1	113	0.00
72	Anthracene	40.000	41.196	-3.0	112	0.00
73	Carbazole	40.000	40.846	-2.1	111	0.00
74	Di-n-butylphthalate	40.000	41.126	-2.8	110	0.00
75 C	Fluoranthene	40.000	39.561	1.1	107	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
77	Benzydine	40.000	43.472	-8.7	86	0.00
78	Pyrene	40.000	44.990	-12.5	106	0.00
79 S	Terphenyl-d14	80.000	86.641	-8.3	104	0.00
80	Butylbenzylphthalate	40.000	41.533	-3.8	103	0.00
81	Benzo(a)anthracene	40.000	40.920	-2.3	106	0.00
82	3,3'-Dichlorobenzidine	40.000	41.018	-2.5	113	0.00
83	Chrysene	40.000	40.906	-2.3	107	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.547	-1.4	105	0.00
85 c	Di-n-octyl phthalate	40.000	41.470	-3.7	107	0.00
86 I	Perylene-d12	20.000	20.000	0.0	113	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	44.324	-10.8	109	0.00
88	Benzo(b)fluoranthene	40.000	38.813	3.0	111	0.00
89	Benzo(k)fluoranthene	40.000	41.650	-4.1	113	0.00
90 C	Benzo(a)pyrene	40.000	41.411	-3.5	111	0.00
91	Dibenzo(a,h)anthracene	40.000	44.439	-11.1	113	0.00
92	Benzo(g,h,i)perylene	40.000	45.248	-13.1	113	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF091624\
Data File : BF139377.D
Acq On : 16 Sep 2024 10:36
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Sep 16 11:10:45 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF091324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 14 02:58:23 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0