

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF091324

Review By	yogesh	Review On	9/16/2024 1:24:18 AM
Supervise By	mohammad	Supervise On	9/16/2024 2:17:37 AM
SubDirectory	BF091324	HP Acquire Method	BNA_F
		HP Processing Method	bf091324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588
CCC	SP6591
Internal Standard/PEM	S12314,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF139344.D	13 Sep 2024 14:04		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF139345.D	13 Sep 2024 14:32		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF139346.D	13 Sep 2024 15:00	Compound #54 removed from 5ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF139347.D	13 Sep 2024 15:28		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF139348.D	13 Sep 2024 15:57	Calibration Failed for Com#77(Benzidine) in both 8270 DOD and NON-DOD	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF139349.D	13 Sep 2024 16:53	The Calibration is Good For 8270 DOD Except com#77 and good for 625.1 Method Except com#77	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF139350.D	13 Sep 2024 17:51	Calibration is Passing for 625.1 Method for all Compounds Except com#77	RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BF139351.D	13 Sep 2024 18:48		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF139352.D	13 Sep 2024 19:44	Compound #6,9 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBF091324	BF139353.D	13 Sep 2024 20:41	Com#77 is failing in ICV for Both 8270 DOD and NON-DOD	RC/JU	Ok
11	PB163259BL	PB163259BL	BF139354.D	13 Sep 2024 21:09		RC/JU	Ok
12	SP6630	SP6630	BF139355.D	13 Sep 2024 21:38	8270 SPIKE	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF091624

Review By	yogesh	Review On	9/17/2024 1:35:51 AM
Supervise By	mohammad	Supervise On	9/19/2024 4:59:59 AM
SubDirectory	BF091624	HP Acquire Method	BNA_F
		HP Processing Method	bf091324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588
CCC	SP6591
Internal Standard/PEM	S12315,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF139376.D	16 Sep 2024 10:09		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF139377.D	16 Sep 2024 10:36		RC/JU	Ok
3	PB163318BL	PB163318BL	BF139378.D	16 Sep 2024 11:04	Analyzed for contamination check, Method is Fail for Com#77	RC/JU	Ok
4	P3845-06	PT-ACIDS-WP	BF139379.D	16 Sep 2024 12:35	PT Sample-Need 5X Dilution.	RC/JU	Dilution
5	P3845-06DL	PT-ACIDS-WPDL	BF139380.D	16 Sep 2024 13:13		RC/JU	Ok
6	P3929-03	72-11930	BF139381.D	16 Sep 2024 16:30		RC/JU	Ok,M
7	P3929-01	BIN-2	BF139382.D	16 Sep 2024 16:58		RC/JU	Ok,M
8	P3947-01	RW-1	BF139383.D	16 Sep 2024 17:27		RC/JU	Ok,M
9	P3947-01MS	RW-1MS	BF139384.D	16 Sep 2024 17:55		RC/JU	Ok,M
10	P3947-01MSD	RW-1MSD	BF139385.D	16 Sep 2024 18:23		RC/JU	Ok
11	P3930-01	DN-B-33	BF139386.D	16 Sep 2024 18:51		RC/JU	Ok
12	P3930-02RE	DN-B-34RE	BF139387.D	16 Sep 2024 19:20	Surrogate Fail	RC/JU	Confirms
13	P3962-01	TR-04-091124	BF139388.D	16 Sep 2024 19:48	Internal Standard Fail	RC/JU	Ok,M
14	P3976-01	RW-2	BF139389.D	16 Sep 2024 20:16		RC/JU	Ok,M
15	P3979-01	EO-03-091224	BF139390.D	16 Sep 2024 20:44		RC/JU	Ok,M
16	P3980-01	HD-02-091224	BF139391.D	16 Sep 2024 21:12	Need further 5X Dilution.	RC/JU	Dilution

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM091424

Review By	yogesh	Review On	9/16/2024 1:25:35 AM
Supervise By	mohammad	Supervise On	9/17/2024 2:00:30 AM
SubDirectory	BM091424	HP Acquire Method	BNA_M
		HP Processing Method	BM091424

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6621,SP6622,SP6623,SP6624,SP6625,SP6626,SP6627,SP6628
CCC	SP6624
Internal Standard/PEM	S12315,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM047519.D	14 Sep 2024 11:17		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM047520.D	14 Sep 2024 11:57		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM047521.D	14 Sep 2024 12:36	Compound #32,54,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BM047522.D	14 Sep 2024 13:16		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM047523.D	14 Sep 2024 13:55	Compound #32,65 Kept on LR and Compound #54 Kept on QR	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BM047524.D	14 Sep 2024 14:35	The Calibration is Good For 8270 DOD and good for 625.1 Method	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BM047525.D	14 Sep 2024 15:14		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BM047526.D	14 Sep 2024 15:54		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BM047527.D	14 Sep 2024 16:34	Compound #77 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBM091424	BM047528.D	14 Sep 2024 17:53		RC/JU	Ok,M
11	PB163160BL	PB163160BL	BM047529.D	14 Sep 2024 18:32		RC/JU	Ok
12	PB163160BS	PB163160BS	BM047530.D	14 Sep 2024 19:12		RC/JU	Ok,M
13	PB163160BSD	PB163160BSD	BM047531.D	14 Sep 2024 19:51	Double Internal Standard added	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM091924

Review By	Jagrut	Review On	9/20/2024 4:22:41 PM
Supervise By	mohammad	Supervise On	9/22/2024 12:55:48 AM
SubDirectory	BM091924	HP Acquire Method	BNA_M
		HP Processing Method	BM091424

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6621,SP6622,SP6623,SP6624,SP6625,SP6626,SP6627,SP6628
CCC	SP6624
Internal Standard/PEM	S12315,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM047565.D	19 Sep 2024 08:54		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM047566.D	19 Sep 2024 11:49		RC/JU	Ok,M
3	PB163318BL	PB163318BL	BM047567.D	19 Sep 2024 12:35		RC/JU	Ok
4	PB163318BS	PB163318BS	BM047568.D	19 Sep 2024 13:15		RC/JU	Ok,M
5	PB163318BSD	PB163318BSD	BM047569.D	19 Sep 2024 13:54		RC/JU	Ok,M
6	P3845-22	RR-TRIAZINE-WP	BM047570.D	19 Sep 2024 14:34	PT Sample, Internal Standard Fail	RC/JU	Not Ok
7	PB163344BS	PB163344BS	BM047571.D	19 Sep 2024 15:14		RC/JU	Ok,M
8	PB163344BSD	PB163344BSD	BM047572.D	19 Sep 2024 15:53		RC/JU	Ok,M
9	PB163344BL	PB163344BL	BM047573.D	19 Sep 2024 16:33		RC/JU	Ok
10	P3956-02	EFF-WW	BM047574.D	19 Sep 2024 17:13	Surrogate Fail	RC/JU	Ok
11	PB163525BL	PB163525BL	BM047575.D	19 Sep 2024 17:53	Internal Standard and Surrogate fail	RC/JU	Not Ok
12	PB163525BS	PB163525BS	BM047576.D	19 Sep 2024 18:32		RC/JU	Ok,M
13	P4103-13	WC-14-7.5-10B	BM047577.D	19 Sep 2024 19:12		RC/JU	Ok
14	P4103-13MS	WC-14-7.5-10BMS	BM047578.D	19 Sep 2024 19:52		RC/JU	Ok,M
15	P4103-13MSD	WC-14-7.5-10BMSD	BM047579.D	19 Sep 2024 20:32		RC/JU	Ok,M

M : Manual Integration