

Surrogate Summary

SW-846

SDG No.: P3845

Client: Chemtech Consulting Group

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P3845-05	PT-BN-WP	2-Methylnaphthalene-d10	0.4	0	0	*	20	139
		Fluoranthene-d10	0.4	0	0	*	30	150
		Nitrobenzene-d5	0.4	101	25264	*	27	123
		2-Fluorobiphenyl	0.4	75.2	18800	*	34	132
		Terphenyl-d14	0.4	84.9	21231	*	35	157
P3845-05DL	PT-BN-WPDL	2-Methylnaphthalene-d10	0.4	0	0	*	20	139
		Fluoranthene-d10	0.4	0	0	*	30	150
		Nitrobenzene-d5	0.4	101	25200	*	27	123
		2-Fluorobiphenyl	0.4	101	25313	*	34	132
		Terphenyl-d14	0.4	127	31800	*	35	157
PB163341BL	PB163341BL	2-Methylnaphthalene-d10	0.4	0.32	79		20	139
		Fluoranthene-d10	0.4	0.37	92		30	150
		Nitrobenzene-d5	0.4	0.30	75		27	123
		2-Fluorobiphenyl	0.4	0.35	86		34	132
		Terphenyl-d14	0.4	0.44	109		35	157
PB163341BS	PB163341BS	2-Methylnaphthalene-d10	0.4	0.39	97		20	139
		Fluoranthene-d10	0.4	0.55	137		30	150
		Nitrobenzene-d5	0.4	0.39	97		27	123
		2-Fluorobiphenyl	0.4	0.41	103		34	132
		Terphenyl-d14	0.4	0.55	137		35	157
PB163341BSD	PB163341BSD	2-Methylnaphthalene-d10	0.4	0.36	89		20	139
		Fluoranthene-d10	0.4	0.37	92		30	150
		Nitrobenzene-d5	0.4	0.36	90		27	123
		2-Fluorobiphenyl	0.4	0.50	126		34	132
		Terphenyl-d14	0.4	0.38	94		35	157