

6C

SEMIVOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: P3845 SAS No.: P3845 SDG No.: P3845
Instrument ID: BNA_N Calibration Date(s): 09/18/2024 09/18/2024
Calibration Time(s): 11:53 15:30

LAB FILE ID:		RRF0.1 = BN034033.D		RRF0.2 = BN034034.D		RRF0.4 = BN034035.D		
		RRF0.8 = BN034036.D		RRF1.6 = BN034037.D		RRF3.2 = BN034038.D		
COMPOUND	RRF0.1	RRF0.2	RRF0.4	RRF0.8	RRF1.6	RRF3.2	RRF	% RSD
2-Methylnaphthalene-d10	0.568	0.554	0.560	0.631	0.627	0.585	0.591	5.4
Fluoranthene-d10	0.965	0.947	0.953	1.059	1.054	1.013	1.009	5.3
2-Fluorophenol	1.201	1.157	1.069	1.203	1.148	1.080	1.135	5.0
Phenol-d6	1.239	1.215	1.189	1.406	1.401	1.371	1.320	7.7
Nitrobenzene-d5	0.270	0.269	0.280	0.331	0.336	0.321	0.306	10.3
2-Fluorobiphenyl	1.629	1.576	1.574	1.722	1.693	1.597	1.627	3.6
4,6-Dinitro-2-methylphenol		0.042	0.046	0.061	0.065	0.072	0.057	21.9
2,4,6-Tribromophenol	0.162	0.156	0.155	0.187	0.192	0.197	0.181	13.5
Pentachlorophenol	0.078	0.065	0.070	0.088	0.095	0.105	0.083	18.3
Terphenyl-d14	0.810	0.767	0.794	0.842	0.821	0.767	0.799	3.5